



Republic of the Philippines
PROVINCE OF PAMPANGA
City of San Fernando, Pampanga

PHILIPPINE BIDDING DOCUMENTS

(As Harmonized with Development Partners)

PROCUREMENT OF GOODS

Government of the Republic of the Philippines

PR No. 25 - 0977

Supply & Delivery of MEDICAL EQUIPMENT AND SUPPLIES for Pampanga Provincial Hospital Clark

Sixth Edition
July 2020

Preface

These Philippine Bidding Documents (PBDs) for the procurement of Goods through Competitive Bidding have been prepared by the Government of the Philippines for use by any branch, constitutional commission or office, agency, department, bureau, office, or instrumentality of the Government of the Philippines, National Government Agencies, including Government-Owned and/or Controlled Corporations, Government Financing Institutions, State Universities and Colleges, and Local Government Units. The procedures and practices presented in this document have been developed through broad experience, and are for mandatory use in projects that are financed in whole or in part by the Government of the Philippines or any foreign government/foreign or international financing institution in accordance with the provisions of the 2016 revised Implementing Rules and Regulations of Republic Act No. 9184.

The Bidding Documents clearly and adequately define, among others: (i) the objectives, scope, and expected outputs and/or results of the proposed contract or Framework Agreement, as the case may be; (ii) the eligibility requirements of Bidders; (iii) the expected contract or Framework Agreement duration, the estimated quantity in the case of procurement of goods, delivery schedule and/or time frame; and (iv) the obligations, duties, and/or functions of the winning bidder.

Glossary of Acronyms, Terms, and Abbreviations

Whenever the following terms or acronyms are used in these bidding documents, they shall have the following respective meanings:

ABC – Approved Budget for the Contract.

BAC – Bids and Awards Committee.

Bid – A signed offer or proposal to undertake a contract submitted by a bidder in response to and in consonance with the requirements of the bidding documents. Also referred to as *Proposal* and *Tender*. (2016 revised IRR, Section 5[c])

Bidder – Refers to a contractor, manufacturer, supplier, distributor and/or consultant who submits a bid in response to the requirements of the Bidding Documents. (2016 revised IRR, Section 5[d])

Bidding Documents – The documents issued by the Procuring Entity as the bases for bids, furnishing all information necessary for a prospective bidder to prepare a bid for the Goods, Infrastructure Projects, and/or Consulting Services required by the Procuring Entity. (2016 revised IRR, Section 5[e])

BDS – Bid Data Sheet

BIR – Bureau of Internal Revenue.

BSP – Bangko Sentral ng Pilipinas.

Consulting Services – Refer to services for Infrastructure Projects and other types of projects or activities of the GOP requiring adequate external technical and professional expertise that are beyond the capability and/or capacity of the GOP to undertake such as, but not limited to: (i) advisory and review services; (ii) pre-investment or feasibility studies; (iii) design; (iv) construction supervision; (v) management and related services; and (vi) other technical services or special studies. (2016 revised IRR, Section 5[i])

CDA - Cooperative Development Authority.

Contract – Refers to the agreement entered into between the Procuring Entity and the Supplier or Manufacturer or Distributor or Service Provider for procurement of Goods and Services; Contractor for Procurement of Infrastructure Projects; or Consultant or Consulting Firm for Procurement of Consulting Services; as the case may be, as recorded in the Contract Form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.

CIF – Cost Insurance and Freight.

CIP – Carriage and Insurance Paid.

CPI – Consumer Price Index.

DDP – Refers to the quoted price of the Goods, which means “delivered duty paid.”

DTI – Department of Trade and Industry.

EXW – Ex works.

FCA – “Free Carrier” shipping point.

FOB – “Free on Board” shipping point.

Foreign-funded Procurement or Foreign-Assisted Project– Refers to procurement whose funding source is from a foreign government, foreign or international financing institution as specified in the Treaty or International or Executive Agreement. (2016 revised IRR, Section 5[b]).

Framework Agreement – Refers to a written agreement between a procuring entity and a supplier or service provider that identifies the terms and conditions, under which specific purchases, otherwise known as “Call-Offs,” are made for the duration of the agreement. It is in the nature of an option contract between the procuring entity and the bidder(s) granting the procuring entity the option to either place an order for any of the goods or services identified in the Framework Agreement List or not buy at all, within a minimum period of one (1) year to a maximum period of three (3) years. (GPPB Resolution No. 27-2019)

GFI – Government Financial Institution.

GOCC – Government-owned and/or –controlled corporation.

Goods – Refer to all items, supplies, materials and general support services, except Consulting Services and Infrastructure Projects, which may be needed in the transaction of public businesses or in the pursuit of any government undertaking, project or activity, whether in the nature of equipment, furniture, stationery, materials for construction, or personal property of any kind, including non-personal or contractual services such as the repair and maintenance of equipment and furniture, as well as trucking, hauling, janitorial, security, and related or analogous services, as well as procurement of materials and supplies provided by the Procuring Entity for such services. The term “related” or “analogous services” shall include, but is not limited to, lease or purchase of office space, media advertisements, health

maintenance services, and other services essential to the operation of the Procuring Entity. (2016 revised IRR, Section 5[r])

GOP – Government of the Philippines.

GPPB – Government Procurement Policy Board.

INCOTERMS – International Commercial Terms.

Infrastructure Projects – Include the construction, improvement, rehabilitation, demolition, repair, restoration or maintenance of roads and bridges, railways, airports, seaports, communication facilities, civil works components of information technology projects, irrigation, flood control and drainage, water supply, sanitation, sewerage and solid waste management systems, shore protection, energy/power and electrification facilities, national buildings, school buildings, hospital buildings, and other related construction projects of the government. Also referred to as *civil works or works*. (2016 revised IRR, Section 5[u])

IB – Invitation to Bid

ITB – Instructions to Bidder

LGUs – Local Government Units.

NFCC – Net Financial Contracting Capacity.

NGA – National Government Agency.

PhilGEPS - Philippine Government Electronic Procurement System.

Procurement Project – refers to a specific or identified procurement covering goods, infrastructure project or consulting services. A Procurement Project shall be described, detailed, and scheduled in the Project Procurement Management Plan prepared by the agency which shall be consolidated in the procuring entity's Annual Procurement Plan. (GPPB Circular No. 06-2019 dated 17 July 2019)

PSA – Philippine Statistics Authority.

SEC – Securities and Exchange Commission.

SLCC – Single Largest Completed Contract.

Supplier – refers to a citizen, or any corporate body or commercial company duly organized and registered under the laws where it is established, habitually established in business and engaged in the manufacture or sale of the merchandise or performance of the general services covered by his bid. (Item 3.8 of GPPB Resolution No. 13-2019, dated 23 May 2019). Supplier as used in these Bidding Documents may likewise refer to a distributor, manufacturer, contractor, or consultant.

UN – United Nations Section

TABLE OF CONTENTS

<i>SECTION I</i>	INVITATION TO BID
<i>SECTION II</i>	INSTRUCTION TO BIDDERS
<i>SECTION III</i>	BID DATA SHEET
<i>SECTION IV</i>	GENERAL CONDITIONS OF CONTRACT
<i>SECTION V</i>	SPECIAL CONDITIONS OF CONTRACT
<i>SECTION VI</i>	SCHEDULE OF REQUIREMENTS
<i>SECTION VII</i>	TECHNICAL SPECIFICATIONS
<i>SECTION VIII</i>	CHECKLIST OF TECHNICAL & FINANCIAL DOCUMENTS
<i>SECTION IX</i>	OTHER BIDDING FORMS

SECTION I. INVITATION TO BID

Republic of the Philippines
PROVINCE OF PAMPANGA
City of San Fernando
INVITATION TO BID NO. 2025-32
MEDIC- BAC

INVITATION TO BID FOR PURCHASE REQUEST NO. 25-0977

The **Provincial Government of Pampanga ("Province")** intends to apply the following sums from the following funds, as follows:

SOURCE OF FUNDING/YEAR	APPROVED BUDGET FOR THE CONTRACT (ABC)	PURCHASE REQUEST NUMBER	DESCRIPTION
General Fund	115,497,530.00	#25-0977	Supply and Delivery of Medical Equipment and Supplies for Pampanga Provincial Hospital Clark

Bids received in excess of the ABC shall be automatically rejected at bid opening.

1. The **Province** invites bids for the above Procurement Project. Delivery of Goods shall be in accordance with the delivery schedule under Section VI. **Schedule of Requirements.** Bidders should have completed, within two years from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II. Instruction to Bidders.
2. Bidding will be conducted through open competitive bidding procedures using non-discretionary "pass/fail" criterion as specified in the 2016 revised Implementing Rules and Regulations (IRR) of Republic Act (RA) No. 9184.

Bidding is open to all interested bidders, whether local or foreign, subject to the conditions for eligibility provided in the 2016 revised IRR of RA No. 9184.

3. Interested bidders may obtain further information from the **Province** and inspect the Bidding Documents at the **Bids and Awards Committee (BAC) Secretariat Office, Ground Floor, Provincial Capitol, City of San Fernando, Pampanga between 8:00 A.M. to 5:00 P.M.**
4. A complete set of Bidding Documents may be purchased by interested Bidders from **February 27, 2025 (Thursday) between 8:00 A.M. to 5:00 P.M. up to March 19, 2025 (Wednesday) until 8:30 A.M. only** from the given address above and upon payment of the applicable fee for the Bidding Documents, pursuant to the latest Guidelines issued by the GPPB, in the following amounts:

PURCHASE REQUEST NUMBER	ABC	AMOUNT
#25-0977	115,497,530.00	50,000.00

The procuring entity shall allow the bidder to present its proof of payment of the fees in person.

5. The **Province** will hold a Pre-Bid Conference on **March 7, 2025 (Friday), 9:00 A.M.** at the **Executive House, Second Floor, Provincial Capitol, City of San Fernando, Pampanga**, which shall be open to prospective bidders.
6. Bids must be duly received by the **Bids and Awards Committee (BAC) Secretariat Office** thru manual submission at **Ground Floor, Provincial Capitol, City of San Fernando, Pampanga on or before 8:45 A.M. of March 19, 2025 (Wednesday).** Late bids shall not be accepted

7. All bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in **ITB Clause 14**.
8. Opening of bids will be on **March 19, 2025 (Wednesday), 9:00 A.M.** at the **Executive House, Second Floor, Provincial Capitol, City of San Fernando, Pampanga**. Bids will be opened in the presence of the bidders' representatives who choose to attend at the activity.
9. Online bid submission is not applicable.
10. The **Province** reserves the right to accept or reject any and all bids, declare a failure bidding, or not award the contract at any time prior to contract award in accordance with Sections 35.6 and 41 of the 2016 revised Implementing Rules and Regulations (IRR) of RA No. 9184, without hereby incurring any liability to the affected bidder or bidders.
11. For further information, please refer to:
FILIPINA R. SEVILLA
Chairperson - Bids and Awards Committee (BAC) Secretariat BAC Secretariat Office, Ground Floor, Provincial Capitol, City of San Fernando, Pampanga
12. **You may visit the following websites:**

For downloading of Bidding Documents: www.pampanga.gov.ph

February 26, 2025

FRANCIS V. MASLOG (SGD)
Chairman
Bids and Awards Committee

SECTION II.
INSTRUCTION TO
BIDDERS

1. Scope of Bid

The Procuring Entity, *The Provincial Government of Pampanga* wishes to receive Bids for the Procurement of **Supply & Delivery of MEDICAL EQUIPMENT AND SUPPLIES for Pampanga Provincial Hospital Clark** with *Identification Number: Purchase Request No. 25 – 0977*.

The Procurement of **Supply & Delivery of MEDICAL EQUIPMENT AND SUPPLIES for Pampanga Provincial Hospital Clark** is composed of **One (1) Lot**, the details of which are described in **Section VII (Technical Specifications)**.

[Note: The Project Identification Number is assigned by the Procuring Entity based on its own coding scheme and is not the same as the PhilGEPS reference number, which is generated after the posting of the bid opportunity on the PhilGEPS website.]

2. Funding Information

2.1. The Provincial Government of Pampanga through the source of funding as indicated below for **General Fund** in the amount of **One Hundred Million Pesos (Php 100,000,000.00)** and **Trust Fund** in the amount of **Fifteen Million Four Hundred Ninety-Seven Thousand Five Hundred Thirty Pesos (Php 15,497,530.00)**.

2.2. The source of funding is **LGU's proposed Local Government Expenditure Program**

3. Bidding Requirements

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or **Invitation to Bid (IB)** by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex "I" of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

5. Eligible Bidders

5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.

5.2.

a. Foreign ownership exceeding those allowed under the rules may participate pursuant to:

i. **When a Treaty or International or Executive Agreement as provided in Section 4 of the RA No. 9184 and its 2016 revised IRR allow foreign bidders to participate;**

b. Foreign ownership limited to those allowed under the rules may participate in this Project.

5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder have an SLCC that is at least one (1) contract similar to the Project the value of which, adjusted to current prices using the PSA's CPI, must be at least equivalent to:

a. **For the procurement of Non-expendable Supplies: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.**

5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under **Instruction to Bidders (ITB)** Clause 18.

7. Subcontracts

The Procuring Entity has prescribed that ***Subcontracting is not allowed.***

8. Pre-Bid Conference

The Procuring Entity will hold a pre-bid conference for this Project on the specified date and time and either at its physical address (***Please refer to Paragraph 6 of Invitation to Bid (IB)***)

9. Clarification and Amendment of Bidding Documents

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the **Invitation to Bid (IB)**, at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.

10. Documents comprising the Bid: Eligibility and Technical Components

10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.

- 10.2. The Bidder's SLCC as indicated in **Instruction to Bidders (ITB)** Clause 5.3 should have been completed within *(Please refer to paragraph 2 of the Invitation to Bid)* prior to the deadline for the submission and receipt of bids.
- 10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated through an apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

11. Documents comprising the Bid: Financial Component

- 11.1. The second bid envelope shall contain the financial documents for the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.
- 11.3. Any bid exceeding the ABC indicated in *paragraph 1 of Invitation to Bid (IB)* shall not be accepted.

12. Bid Prices

- 12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:
 - a. For Goods offered from within the Procuring Entity's country:
 - i. The price of the Goods quoted EXW (ex-works, ex-factory, ex-warehouse, ex-showroom, or off-the-shelf, as applicable)
 - ii. The cost of all customs duties and sales and other taxes already paid or payable;
 - iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
 - iv. The price of other (incidental) services, if any, listed in the BDS.
 - b. For Goods offered from abroad:
 - i. Unless otherwise stated in the **Bid Data Sheet (BDS)**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.

- ii. The price of other (incidental) services, if any, as listed in **Section VII (Technical Specifications)**.

13. Bid and Payment Currencies

- 13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.
- 13.2. Payment of the contract price shall be made in **PHILIPPINE PESO**.

14. Security

- 14.1. The Bidder shall submit a Bid Securing Declaration¹ or any form of Bid Security in the amount indicated in the **Bid Data Sheet (BDS)**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.
- 14.2. The Bid and Bid Security shall be valid until 120 calendar days from date of the Opening of Bids. Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.

15. Sealing and Marking of Bids

Each Bidder shall submit one copy of the first and second components of its Bid.

The Procuring Entity may request additional hard copies and/or electronic copies of the Bid. However, failure of the Bidders to comply with the said request shall not be a ground for disqualification.

If the Procuring Entity allows the submission of bids through online submission or any other electronic means, the Bidder shall submit an electronic copy of its Bid, which must be digitally signed. An electronic copy that cannot be opened or is corrupted shall be considered non-responsive and, thus, automatically disqualified.

16. Deadline for Submission of Bids

- 16.1. The Bidders shall submit on the specified date and time and either at physical address or through online submission as indicated in paragraph 7 of the **IB**.

17. Opening and Preliminary Examination of Bids

- 17.1. The BAC shall open the Bids in public at the time, on the date, and at the place specified *in paragraph 5 of the Invitation to Bid (IB)*. The Bidders' representatives who are present shall sign a register evidencing their attendance. In case videoconferencing, webcasting or other similar technologies will be used, attendance of participants shall likewise be recorded by the BAC Secretariat.

¹ In the case of Framework Agreement, the undertaking shall refer to entering into contract with the Procuring Entity and furnishing of the performance security or the performance securing declaration within ten (10) calendar days from receipt of Notice to Execute Framework Agreement.

In case the Bids cannot be opened as scheduled due to justifiable reasons, the rescheduling requirements under Section 29 of the 2016 revised IRR of RA No. 9184 shall prevail.

- 17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

18. Domestic Preference

- 18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

19. Detailed Evaluation and Comparison of Bids

- 19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated “*passed*,” using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.
- 19.2. If the Project allows partial bids under **Section I of the ITB**, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by **Instruction to Bidder (ITB)** Clause 15 shall be submitted for each lot or item separately.
- 19.3. The descriptions of the lots or items shall be indicated in **Section VII (Technical Specifications)**, although the ABCs of these lots or items are indicated in the **Bid Data Sheet (BDS)** for purposes of the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.
- 19.4. The Project shall be awarded as ***One (1) Project having several items that shall be awarded as one (1) contract.***
- 19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

20. Post-Qualification

Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid, the bidder shall submit its latest income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the **BDS**;

21. Signing of the Contract

The documents required in Section 37.2 of the 2016 revised IRR of RA No. 9184 shall form part of the Contract. Additional Contract documents are indicated in the **BDS**.

SECTION III.
BID DATA SHEET

Bid Data Sheet

ITB Clause	
5.3	<p>For this purpose, contracts similar to the Project shall be:</p> <p style="padding-left: 40px;">a. <i>“Medical Equipment/Supplies”</i></p> <p style="padding-left: 40px;">b. Completed within <i>two (2) years</i> prior to the deadline for the submission and receipt of bids.</p>
7	<i>“Subcontracting is not Allowed”</i>
13.2(a)	The Bid prices for Goods supplied shall be quoted in <i>Philippine Peso</i>
14.1	<p>The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms:</p> <ol style="list-style-type: none"> 1. <i>Php 2,309,950.60</i> 2% of ABC, if bid security is in cash, cashier’s/manager’s check, bank draft/guarantee or irrevocable letter of credit; 2. <i>Php 5,774,876.50</i> 5% of ABC, if bid security is in Surety bond; or <p>Any combination of the foregoing proportionate to the share of form with respect to total amount of security</p>
15	<p>Each Bidder shall submit One (1) Copy of the Technical and Financial components of its bid. The two components shall be placed into separate envelope forming the Two (2) Envelopes which shall be marked as follows:</p> <p style="padding-left: 80px;">First Envelope – Technical Components Second Envelope – Financial Components</p> <p>The Two Envelopes shall be enclosed in a single envelope referred to as the “Mother Envelope” which shall be marked as follows:</p> <div style="border: 1px solid black; padding: 10px; margin: 10px auto; width: 80%;"> <p style="text-align: center;">ENVELOPE MARKING (MOTHER ENVELOPE)</p> <p style="text-align: center;">Name and Address of Bidder</p> <p style="text-align: center;">The Chairperson Bids and Awards Committee Provincial Government of Pampanga Capitol Compound, CSFP</p> <p style="text-align: center;">Name of Project Contract ID</p> <p style="text-align: center;">DO NOT OPEN BEFORE: [Insert date & time of opening of bids]</p> </div> <p>All envelopes should be properly sealed, signed and labelled.</p>
19.3	Partial bid is not allowed. The goods are grouped in a single lot and the lot shall not be divided into sub-lots for the purpose of bidding, evaluation, and contract award.
20	<p><i>The bidder with the Lowest Calculated Bid shall submit ALL of the following post qualification requirements:</i></p> <ol style="list-style-type: none"> 1. <i>Copy of the following valid and current Eligibility and Financial Documents listed in PhilGEPS Certificate of Registration (Platinum Membership), as follows:</i> <ol style="list-style-type: none"> a) <i>Registration Certificate issued by Security and Exchange Commission (SEC) or Department of Trade and Industry (DTI);</i> b) <i>Mayor’s Permit or its equivalent document;</i>

	<p>c) Tax Clearance; and d) Latest Audited Financial Statement</p> <p>2. License to Operate as Medical Device Manufacturer/Trader/Distributor issued by FDA.</p> <p>3. Must have authorized Dealership/Sub Dealership/Distributor/Sub Distributor for items 1, 10, 11, 24, 25, 26, 27, 28, 43, 44, 58, 59, 61, 72, 82, 84, 120, 137, 142, 143, 144, 145, 146, 147, 148, 157, 166, 167, 195, 200, 201, and 228.</p> <p>4. Must have Certificate of Medical Device Notification from FDA for items 1, 28, 120, 137, 195, and 228.</p>
21	No further instruction.

**SECTION IV.
GENERAL CONDITIONS
OF CONTRACT**

1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the **Special Conditions of Contract (SCC)**.

2. Advance Payment and Terms of Payment

- 2.1. Advance payment of the contract amount is provided under Annex "D" of the revised 2016 IRR of RA No. 9184.
- 2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the **SCC**.

3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184.

4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project Specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual and, if applicable, additional tests specified in the **SCC, Section VII (Technical Specifications)**. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

5. Warranty

- 5.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184.

5.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

6. Liability of the Supplier

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity.

SECTION V.
SPECIAL CONDITIONS OF
CONTRACT

Special Conditions of Contract

GCC Clause	
1	<p>Delivery and Documents –</p> <p>For purposes of the Contract, “EXW,” “FOB,” “FCA,” “CIF,” “CIP,” “DDP” and other trade terms used to describe the obligations of the parties shall have the meanings assigned to them by the current edition of INCOTERMS published by the International Chamber of Commerce, Paris. The Delivery terms of this Contract shall be as follows:</p> <p><i>[For Goods supplied from abroad, state:]</i> “The delivery terms applicable to the Contract are DDP delivered to (<i>refer to Section VI. Schedule of Requirements</i>) in accordance with INCOTERMS.”</p> <p><i>[For Goods supplied from within the Philippines, state:]</i> “The delivery terms applicable to this Contract (<i>refer to Section VI. Schedule of Requirements</i>). Risk and title will pass from the Supplier to the Procuring Entity upon receipt and final acceptance of the Goods at their final destination.”</p> <p>Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements).</p> <p>For purposes of this Clause the Procuring Entity’s Representative at the Project Site is as stated above.</p> <p>Incidental Services –</p> <p>The Supplier is required to provide all of the following services, including additional services, if any, specified in Section VI. Schedule of Requirements:</p> <p>The Contract price for the Goods shall include the prices charged by the Supplier for incidental services and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.</p>
3	<p>Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184.</p> <p style="text-align: center;">(a) The Procuring Entity’s address for Notices is: <i>BIDS AND AWARDS COMMITTEE – SECRETARIAT OFFICE, Ground Floor, Provincial Capitol Bldg., City of San Fernando, Pampanga</i></p> <p>The Supplier’s address for Notices is: _____</p>
5.1	<p><i>With two (2) years warranty on parts and service for items number 17-19,32,33,37,52,55,56,58,60,68,69,75,80-83,85,86,103-105,113,123,154,157,162,167,196,201,205,218,221-223;</i></p> <p><i>With three (3) years warranty on parts and service and two (2) years extended warranty on service for items number 1,10,11,24-26,28,29,31,43,44,59,61,72,84,120,137,142-149,166 and 202.</i></p>
5.2	<p><i>Replacement of defective items within five (5) calendar days upon notification by the end user.</i></p>

SECTION VI.
SCHEDULE OF
REQUIREMENTS

SCHEDULE OF REQUIREMENT

ITEM NO.	Item Description	Quantity	Delivery Date
	Supply & Delivery of MEDICAL EQUIPMENTS AND SUPPLIES for Pampanga Provincial Hospital Clark	1 lot (Please refer to Section VII. Technical Specifications for the quantity of items)	<i>Delivery and /or commissioning within forty-five (45) calendar days upon receipt of NTP at the Pampanga Provincial Hospital Clark</i>

Additional Requirement/Condition:

1. *With at least five (5) Biomedical Engineer with Certificate of Training or Accreditation or Diploma or Completion or License or TESDA Certificate of the Biomedical Equipment Engineer or Technical or Certifier;*
2. *Printed document material with the brand/model/specifications initially for items 1, 10, 11, 17, 19, 24, 25, 26, 27, 28, 32, 33, 43, 44, 58, 59, 61, 69, 72, 82, 84, 120, 137, 142, 143, 144, 145, 146, 147, 148, 157, 166, 167, 195, 200, 201, 204, 221 and 228 must be submitted together with the bid for evaluation;*
3. *With service facility in Pampanga for after sales service;*
4. *With valid inventory equivalent to the quantity requirement of the project for items 2, 3, 7, 9, 12, 14, 17, 24, 25, 26, 27, 43, 44, 58, 59, 61, 70, 72, 76, 81, 82, 83, 84, 85, 86, 96, 98, 100, 117, 118, 119, 121, 122, 123, 128, 129, 142, 143, 144, 147, 148, 150, 153, 163, 166, 169, 178, 179, 188, 189, 190, 193, 196, 197, 200, 204, 212, 214, 215, 218, 219 and 233 to assure product availability;*
5. *With valid inventory at least twice the quantity requirement of the project for items 4, 5, 6, 8, 11, 13, 15, 16, 18, 19, 20, 21, 22, 23, 29 to 42, 45 to 57, 60, 62 to 69, 71, 73, 74, 75, 77 to 80, 87 to 95, 97, 99, 101 to 116, 124 to 127, 130 to 136, 138 to 141, 146, 149, 151, 152, 154 to 156, 158 to 162, 164, 165, 168, 170 to 177, 180 to 187, 191, 192, 194, 198, 199, 202, 203, 205 to 211, 213, 216, 217, 220 to 227, 229 to 232 to assure product availability;*
6. *Shall provide Operations Manual for all Medical Equipment and staff training of hospital personnel and shall issue necessary Training Certificate;*
7. *Calibration Certificate of all calibrated Medical Equipment duly signed by the Biomedical Engineer;*
8. *24/7 on call technical support service within the warranty period;*
9. *All alteration works on the building which are necessary in the installation of the equipment must be restored in its original state by the supplier at no cost to the Procuring Entity;*

10. *Must provide training for basic Echocardiogram with certification as Echocardiogram Technologies for item 1;*
11. *Process FDA licensing for permit to operate item stand along X-ray machine; and*
12. *All Medical Equipment must be installed and/ or commissioned at the Hospital facility.*

I hereby certify/undertake to comply and deliver all the above requirements.

Name of Company/Bidder

Signature Over Printed Name of Representative

Date

SECTION VII.
TECHNICAL
SPECIFICATIONS

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
1	1	unit	<p>2D Echo Machine</p> <p>Specification:</p> <p>Dimensions: Approx. 1840mm (H) x 600mm (W) x 970mm (D)</p> <p>Electrical Power: Voltage: 100-240 VAC Frequency: 50/60Hz</p> <p>Console Design: Active Probe Ports: 4 active ports (standard) Integrated Speakers for audio output Probe Holders: at least 5 holders, including one slot for endocavitary probes Monitor and Touch Panel: Monitor: at least 23.8" size, 1920 x 1080 high-resolution LED backlit monitor Touch Panel: at least 15.6" screen with multi-gesture recognition and customizable layout Tilt Angle: approx. +45° / -90° Rotation Angle: approx. -90° / +90°</p> <p>INCLUSIVE OF INSTALLATION/COMMISSIONING</p>	
2	20	pc	Adson forcep with teeth, 5 1/2", stainless	
3	20	pc	Adson forcep without teeth, 5 1/2", stainless	
4	1	pc	Alcohol lamp, glass, capacity: at least 50mL	
5	2	pc	Allis forcep, (5x6) 6", stainless	
6	4	pc	Amniotome hook, 10 1/4", stainless	
7	20	pc	Army navy retractor, 8 1/4", stainless	
8	2	pc	Aspirating Dental Syringe, 1.8ml	
9	40	pc	Aspirator for Serological Pipette rubber aspirator with plastic tip, standard size	
10	1	unit	<p>Autoclave Machine 500L</p> <p>Specification:</p> <p>Sterilizing chamber volume: at least 500L Working pressure: at least 0.22MPa Working temp: at least 134°C Sterilization Time: approx. 0-60 mins Drying Time: approx. 0-60 mins Heat average: at least $\leq \pm 1^\circ\text{C}$ Power: approx. 18KW/380V 50Hz Dimension: approx. 1800mm x 900mm x 1820mm</p> <p>INCLUSIVE OF INSTALLATION/COMMISSIONING</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
11	2	unit	<p>Automatic External Defibrillator (AED) With Disposable Battery Specification:</p> <p>Adult and Pediatric Modes: Simple switch to change between modes for adult and pediatric use.</p> <p>Energy Capacity: at least 360J energy delivery for defibrillation</p> <p>CPR Metronome: Provides a rhythm for chest compressions Pre-connected Pads: Pads are already attached for quick deployment Pad and Battery Lifecycle: Up to 5 years for both the pads and battery</p>	
12	19	set	<p>Weight: approx. 1-2 kg</p> <p>Bag valve mask resuscitator with rebreather, adult Inclusive of the following:</p> <ul style="list-style-type: none"> 1 Tackle box 1 Ambu bag mask, adult 1 Reservoir Bag 1 Manual Resuscitator 1 Connector/Oxygen tubing, adult 1 Additional accessories (Airway, mouth opener) 	
13	7	set	<p>Bag valve mask resuscitator with rebreather, infant Inclusive of the following:</p> <ul style="list-style-type: none"> 1 Tackle box 1 Ambu bag mask, infant 1 Reservoir Bag 1 Manual Resuscitator 1 Connector/Oxygen tubing, infant 1 Additional accessories (Airway, mouth opener) 	
14	22	set	<p>Bag valve mask resuscitator with rebreather, pedia Inclusive of the following:</p> <ul style="list-style-type: none"> 1 Tackle box 1 Ambu bag mask, pedia 1 Reservoir Bag 1 Manual Resuscitator 1 Connector/Oxygen tubing, pedia 1 Additional accessories (Airway, mouth opener) 	
15	2	set	Bakes dilator set, 9's/set , sizes: 3mm to 11mm, stainless	
16	8	pc	Bandage scissor, 8", stainless	
17	15	set	<p>Bassinet/Infant Bed Specification:</p> <p>Frame: Material: Stainless steel frame Basin Material: Acrylic basin</p> <p>Mattress: Includes a foam mattress for comfort and support.</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
18	4 set		<p>Base: Frame for Storage: Base includes a frame for storing items, providing convenient storage space.</p> <p>Castors: Castors with Brakes: Equipped with castors (wheels) for easy mobility, with brakes to secure the bed in place when needed.</p> <p>Basin Dimensions: approx. 810mm (L) x 460mm (W) x 810mm (H) Mattress Dimensions: approx. 800mm (L) x 450mm (W)</p> <p>Thickness: Standard thickness suitable for infant comfort.</p> <p>Bilateral Tubal Ligation Set 10- Kelly Forcep 7", 4- Allis Forcep 6" 1- Tissue Forcep 5 1/2" 1- Thumb Forcep 5 1/2" 1- Needle Holder 6" 1- Needle Holder 7" 1- Scalpel 5 1/4" 2- Bobcock Forcep 6 1/4" 5- Towel Clips 5 1/2" 2- Army Navy 8 1/4"</p>	
19	2 unit		<p>Bili Light/ Phototherapy Specification: Operating Time: at least 100,000 hours Light Sources: Blue LEDs LED Configuration: 8 Blue LEDs for light emission. Maximum Height: 200 cm maximum height for adjustable positioning</p>	
20	4 pc		Bulldog clamp, 4", curve, stainless	
21	1 pc		<p>Calibrated Hygrometer Thermometer Specification: Type: Digital LCD Thermometer Electronic Temperature Humidity Meter Weather Station: Suitable for both indoor and outdoor use.</p> <p>Display: Display Size: Approx. 2.0-3.9 inches Power Type: 1x AAA Battery</p>	
22	2 pc		Caliper, at least 300mm, stainless	
23	6 pc		Cardiac Board, approx. L-59.5 x W- 41 x H-7cm, plastic	
24	1 unit		<p>Cardiac Monitor with Cuffs and Oximeter for Neonate Specification:</p>	

**Technical Specification
PR No. 25 - 0977**

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
25	19	unit	<p>Display: Screen Size: At least 12.1-inch color touchscreen for clear and easy viewing. Resolution: At least 1280 x 800 pixels for high display clarity. Display Channels: Supports up to 10 waveform channels</p> <p>ECG: Lead Set: Approx. 3 to 5 leads</p> <p>Sweep Speed Options: Adjustable sweep speeds of 6.25 mm/s, 12.5 mm/s, 25 mm/s, and 50 mm/s for flexibility in monitoring</p> <p>Heart Rate Monitoring: HR Range (Pediatric/Neonate): approx. 10 to 350 bpm Cuff Size: Suitable for neonates with a cuff size of approximately 7-13 cm.</p> <p>Cardiac Monitor with Oximeter and Cardiac Cables for Adult Specification:</p>	
26	4	unit	<p>Display: Screen Size: At least 12.1-inch color touchscreen for clear and easy viewing. Resolution: At least 1280 x 800 pixels for high display clarity. Display Channels: Supports up to 10 waveform channels</p> <p>ECG: Lead Set: Approx. 3 to 5 lead</p> <p>Sweep Speed Options: Adjustable sweep speeds of 6.25 mm/s, 12.5 mm/s, 25 mm/s, and 50 mm/s for flexibility in monitoring</p> <p>Heart Rate Monitoring: HR Range (Adult): approx. 10 to 300 bpm Cuff Size: Suitable for adult use with a cuff size of approximately 25-47 cm.</p> <p>Cardiac Monitor with Oximeter and Cardiac Cables for Pedia Specification:</p> <p>Display: Screen Size: At least 12.1-inch color touchscreen for clear and east viewing Resolution: At least 1280 x 800 pixels for high display clarity Display Channels: Supports up to 10 waveform channels</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
27	1	unit	<p>ECG: Lead Set: Approx. 3 to 5 lead Sweep Speed Options: Adjustable sweep speeds of 6.25 mm/s, 12.5 mm/s, 25 mm/s, and 50 mm/s.</p> <p>Heart Rate Monitoring: HR Range (Pediatric/Neonate): Approx. 10 to 350 bpm. Cuff Size: Cuff size for pediatric use, approximately 18-26 cm.</p> <p>Cardiotocography Monitor with cable Specification:</p> <p>Dimensions & Weight: Dimensions: Approx. 19.1 x 18.7 x 20.1 cm Weight: At least 5.0 kg</p> <p>Power: Input: 100-240V for universal compatibility. Dissipation: Battery-powered, with at least 1 hour of battery operation when fully charged at 25°C.</p> <p>Environmental Conditions: Operating Temperature: Approx. 10-40°C. Operating Humidity: Approx. 5-85% non-condensing.</p> <p>Monitoring Features: Doppler Ultrasound Monitoring for FHR and uterine contractions.</p> <p>Parameters: FHR Range: 30-240 BPM Uterine Contraction (UC) Range: 0-99 relative units</p>	
28	1	unit	<p>Paper: Paper Type: Standard paper for recording data. Paper Size: Approx. 130 mm x 120 mm x 20 mm Paper Speeds: 1, 2, and 3 cm/min ± 1%</p> <p>Cautery Machine Specification:</p> <p>Display: Screen Size: At least 15.7 cm (6.2 inches) LCD touchscreen.</p> <p>Duty Cycle: Capable of operating with a 25% duty cycle (at least 10 seconds active and 30 seconds inactive in any mode for up to 4 hours).</p> <p>Dimensions: Height: Approx. 14.6 cm Width: Approx. 33.5 cm</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
29	5 set		<p>Length: Approx. 43.6 cm</p> <p>Functionality: Capable of monopolar and bipolar coagulation. Recall program to retain previous settings. Auto-bipolar function for easy switching between modes. Software upgradeable and Wi-Fi ready for future-proofing.</p> <p>Accessories per Unit: Trolley: Compatible trolley for the unit. Foot Pedals: One monopolar foot pedal and one bipolar foot pedal, both safe medical grade and splash-proof.</p> <p>Cesarean Set 10 Kelly forcep 7" 4 Allis forcep 6" 1 Tissue forcep 5 1/4" 1 Thumb forcep 5 1/4" 1 Debaquey 1 Needle holder 6" 1 Needle holder 7" 1 Scalpel 5 1/4" 2 Bobcock 6 1/4" 5 Towel Clips 5 1/4"</p>	
30	2 pc		<p>Richardson Eastman Retractor (DOUBLE ENDED SET OF TWO 10") 1 Bladder Retractor Chisel Specification:</p>	
31	3 set		<p>Hardness: Approx. 55-57 HRC on the cutting edge. Material: Made from stainless steel Body: Rounded body</p> <p>Cholecystectomy Set 10 Kelly Forcep 7" 2 Allis forcep 7 1/2" 1 Tissue forcep 5 1/2" 1 Thumb forcep 5 1/2" 1 Debaquey forcep 1 Needle holder 6" 1 Needle holder 7" 1 Scalpel 5 1/4" 2 Bobcock 6 1/4" 2 Mixer forcep 7 1/4" 5 Towel clips 5 1/2" 2 Army Navy 9 1/4" 2 Richardson 9 1/2", 3/4 x 1" grip handle 1 Deaver Retractor (Medium) 2x12" 1 Malleable 1x1/2 x 13" 1 Hegar uterine dilator set (double end) 1 Long thumb 8"</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
32	8	unit	<p>Clinical height and weight scale (For Adult & Pedia)</p> <p>Specification:</p> <p>Height Measurement: Height Range: approx. 70 - 190 cm Height Measurement Unit: Displayed in cm or inches (user-selectable).</p> <p>Weight Measurement: Weight Capacity: approx. 2-180 kg Weight Measurement Units: Displayed in kg or lb (user-selectable).</p> <p>Design: Platform: Sturdy and anti-slip surface for patient safety during measurement. Material: High-quality, durable material (stainless steel or strong plastic) Non-slip Feet: For stability on various floor surfaces.</p>	
33	6	unit	<p>Clinical weighing scale (For Adult & Pedia)</p> <p>Specification:</p> <p>Weight Measurement: Weight Capacity: Approx. 2-180 kg</p> <p>Weight Measurement Units: Display in kg or lb (user-selectable).</p> <p>Platform: Platform Size: Sufficiently large to accommodate both adults and children comfortably (approx. 35 cm x 35 cm or larger). Material: High-quality, durable material (stainless steel or strong plastic) Non-slip Feet: For stability on various floor surfaces</p>	
34	10	pack	Cover slip, thick, thickness: approx. 0.13-0.16mm, 200s/pack	
35	2	set	Curette #00, #1, #2, #3 Sharp/Dull, stainless	
36	2	pc	Curved hemostat 6", stainless	
37	2	set	<p>Dilation & Curettage (DC) Set</p> <p>1 Sims uterine sound 12½"</p> <p>1 Sims uterine cuette (SHARP & DULL) #00</p> <p>1 Sims uterine cuette (SHARP & DULL) #1</p> <p>1 Sims uterine cuette (SHARP & DULL) #3</p> <p>1 Sims uterine cuette (SHARP & DULL) #5</p> <p>1 Sims vaginal retractor</p> <p>1 Schoeder Braun Tenaculum 9½"</p> <p>1 Bozeman uterine forcep 9½"</p> <p>1 Foerster sponge forcep 9½"</p> <p>1 Beacham amniotome 9½"</p> <p>1 Heagr uterine dilator set (double end)</p> <p>1 Auvard weighted vaginal speculum 9"</p> <p>1 Graves speculum, small</p> <p>1 Graves speculum, medium</p>	

**Technical Specification
PR No. 25 - 0977**

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
38	5	pc	1 Graves speculum, large	
39	5	pc	1 Allis forcep 9½"	
40	5	pc	Deaver retractor, blade size: 12x1", stainless	
41	5	pc	Deaver Retractor, blade size: 12x2", stainless	
42	5	pc	Deaver Retractor, blade size: 12x3", stainless	
43	10	unit	Debakey forceps, 6", stainless	
			Debakey forceps, 9 1/2", stainless	
			Defibrillator with Paddles	
			SPECIFICATION:	
			Display	
			Screen type: High-resolution display	
			Screen size: at least 7 inches diagonally	
			Sweep speed: Standard sweep speed for ECG monitoring	
			Information: Heart rate, Lead/Pads, Alarm On/Off, SpO2, AED	
			Functions and Prompts,	
			Alarm Selection and Limits, Delivered Energy (for defibrillation)	
			Waveform: Biphasic or equivalent	
			Energy Display: Monitor display indicates both selected and delivered energy.	
			Charge prompt: type: Voice and Visual prompts	
			ECG Monitoring	
			Patient Connection: 3-lead ECG cable, or 5-lead ECG cable, paddles	
			Lead Selection: displays common lead configurations, including at least I,II,III,V	
			ECG size: adjustable ECG scale for different monitoring preferences	
			Heart Rate Alarm: on/off displayed on monitor, user-selectable	
			Smart Alarms: alerts via audio or visual prompts for shockable rhythms	
44	3	unit	Delivery Room Bed Table with Stirrups	
			Specification:	
			Table length mm : approx.1900 with a tolerance of ±50	
			Table width mm : approx. 600mm with a tolerance of ±10	
			Height Adjustment mm : approx. 700mm-900mm	
			Backboard Upwards : at least 50°	
			Backboard Downwards : at least 3°	
			Trendelenburg : adjustable to at least 10°	
			Reverse Trendelenburg Position: adjustable to at least 20°	
			Leg board Outwards : at least 90°	
			Working Area : approx. 600mm x 600mm	
			- INCLUSIVE OF INSTALLATION	
45	2	pc	Dental forcep #150, stainless	
46	2	pc	Dental forcep #151, stainless	
47	2	pc	Dental forcep #17, stainless	
48	2	pc	Dental forcep #18L, stainless	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
49	2	pc	Dental forcep #18 R, stainless	
50	2	pc	Dental forcep #44, stainless	
51	2	pc	Dental forcep #69, stainless	
52	1	unit	Dental Light Curing Machine Specification: Input voltage: approx. 110-220V Wavelength: approx. 420-480nm Illumination: approx. 1200-1500mw/cm ² Includes: 1 x Fiber probe 1 x Charger with Adapter 1 x Light shield	
53	25	pc	Disposable dropper, 5mL	
54	3	pc	Dressing table Specification: Size: approx. 670mm (L) x 450mm (W) x 1,000mm (H) Stainless steel With at least 2 drawers	
55	35	pc	Droplight/Gooseneck Specification: with wheels dimmer switch, 50W warm LED bulb stand made of stainless steel Ergonomically designed: light and portable, with steady base, with flexible neck for ease in use and adjustable height	
56	1	unit	Dental Ultrasound Scaler Specification: Output power: approx. 3W-20W Frequency: approx. 28kHz+3kHz Dimension: approx.188mm x 159mm x 81mm Main unit weight: at least 0.65kg	
57	1	unit	Dry heat sterilizer Specification: Voltage: at least 220V Frequency: approx. 50-60Hz Power Input: at least 300W Capacity: at least 28L Full-Load: 3kg	
58	3	unit	Electrocardiogram (ECG) Machine 3 Channels Specification: Lead: supports 3-lead Processing: ECG analysis sampling rate: at least 500 samples/ second (sps) Digital sampling rate: at least 500 samples/ second/ channel	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
59	1	unit	<p>Resolution: at least 1 μV / LSB @ 500 sps Heart rate meter: approx. 30 to 300 BPM, with accuracy of \pm 10% or \pm5 BPM, whichever is greater Startup time: < 5 seconds</p> <p>Display: Color: 24 bit Color display Display type: approx. 480x320 large high-resolution display Display data: shows Patient ID, gender, age, heart rate, clock, battery power indicator, waveforms, lead labels, speed, gain, filter settings, warning messages, information messages, network status and USB</p> <p>AC Power: Input voltage: approx. 100 to 240VAC \pm 10% Input power: at least 60VA</p> <p>Battery: Battery type: Rechargeable Lithium-ion battery Battery capacity: 11.1 V typical, 2500 mAh, with 6 hours of continuous operation without recording Battery charge time: at least 3 hours with power off</p> <p>Writer: Writer technology: Thermal dot array Writer Width: at least 80mm Writer speed: 5, 12.5, 25, 50 mm/s for printing 3 leads Writer speed accuracy: at least \pm5% Writer amplitude accuracy: at least \pm5 % Writer resolution: Horizontal 32 dots/mm at 25 mm/s, Vertical 8 dots/mm</p> <p>Paper type: Thermal roll paper (at least 20 m)</p> <p>Internal storage: at least 800 ECGs in Internal Memory</p> <p>Accessories: ECG Patient Cable with banana plugs, Limb Clamps, Chest Bulbs Recorder Paper- 1 roll Power Cord</p> <p>Electrocardiogram (ECG) Machine with Leads, 12 Channels Specification: Lead: supports 12 channels</p> <p>Processing: ECG analysis sampling rate: at least 1000 samples/ second (sps)</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
60	6 set		<p>Digital sampling rate: at least 1000 samples/ second/ channel Resolution: at least 1 μV / LSB @ 500 sps Heart rate meter: approx. 30 to 300 BPM, with accuracy of \pm 10% or \pm5 BPM, whichever is greater Startup time: < 5 seconds</p> <p>Display: Color: 24 bit Color display Display type: at least 800x480, large high-resolution display Display data: shows Patient ID, gender, age, heart rate, clock, battery power indicator, waveforms, lead labels, speed, gain, filter settings, warning messages, information messages, network status and USB</p> <p>AC Power: Input voltage: approx. 100 to 240VAC \pm 10% Input power: at least 60VA</p> <p>Battery: Battery type: Rechargeable Lithium-ion battery Battery capacity: 11.1 V typical, 2500 mAh, with 6 hours of continuous operation without recording Battery charge time: at least 4 hours with power off</p> <p>Writer: Writer technology: Thermal dot array Writer Width: at least 80mm Writer speed: 5, 12.5, 25, 50 mm/s Writer speed accuracy: at least \pm5% Writer amplitude accuracy: at least \pm5 % Writer resolution: Horizontal 32 dots/mm at 25 mm/s, Vertical 8 dots/mm</p> <p>Paper type: Thermal roll paper, 215mm*20m roll or 210*140mm folder, high resolution thermal paper</p> <p>Internal storage: at least 800 ECGs in Internal Memory</p> <p>Accessories: ECG Patient Cable with banana plugs, Limb Clamps, Chest Bulbs Recorder Paper- 1 roll Power Cord EENT diagnostic set with Ophthalmoscope with lens wheel Otoscope with swivel magnifying lens and at least 3 sizes ear speculum Nasal speculum Bent arm illuminator Battery handle</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
61	4	unit	2 sizes laryngeal mirrors Tongue depressor holder Electric ICU Bed, at least 5 functionality Specification: Dimension: at least L2140 x W1060 mm + 5% Adjustable height: approx. 390 ~ 760 mm (steel platform)/ 400 ~ 770 mm (PP platform) Back raising: approx. 0 ~ 70° Knee raising: approx. 0 ~ 40° Trendelenburg and rev. trendelenburg: approx.+15° ~ -15° Calf adjustment: at least 10 steps Auto-regression: Back/buttock: at least 10 cm/ 2 cm Control system: Low voltage motor system Recharge battery: Standard (30 times use after full charge) Urinal hook: Standard (both sides with hooks) IV pole and socket: 1 units IV pole with 4 units socket Sleeping surface: 4 sections movable PP platform with ventilation hole and body strap holder Head / foot board: PP material, detachable for emergency Caster: at least 125 mm Locking system: Central locking system Lifting pole and socket: Available both sides of back Loading capacity: Max. 250 kg INCLUSIVE OF INSTALLATION/COMMISSIONING	
62	2	pc	Elevator 301, at least 2.5mm straight tip, stainless	
63	2	pc	Elevator 304, at least 3.5mm straight tip, stainless	
64	8	unit	Emergency Cart Specification: Size: at least 750(L)x475(W)x920mm(H) Defibrillator Shelf: 1pc Sliding Side Shelf: 1pc Sharp Container: 1pc Dust Baskets: at least 2pcs IV Pole: 1pc CPR Board: 1pc	
65	1	kit	Emergency spill kit Content per Kit: PPE set freshly prepared 10% Sodium hypochlorite solution 1L 70% Isopropyl Alcohol, 500ml autoclavable plastic bags (Small, 100s/roll) (Medium, 100s/roll) (Large, 100s/roll) Ordinary forceps Tissue paper/ Kitchen towel, pull ups, 2 ply, 100 sheets/pack Broom & Dustpan, small handheld Antihistamine tablet	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
66	6	pc	Instruction sheet	
67	4	pc	"Do not enter" signage Emergency toolbox, made of aluminum, standard size	
68	9	unit	Examination rotating stool Specification: made of stainless steel at least 5x50mm diameter swivel castors upholstery with PVC covered pneumatic height adjustment height: approx. 559-771mm	
69	2	unit	Examining table Specification: Steel construction Scratch & chip resistant epoxy powder coating Upholstery with PVC cover with PU foam of 75mm thick Ratchet backrest of 5 positions At least 1 cabinet and 2 drawers Dimension (mm): approx. 1830(L) x 610(W) x 760(H)	
70	144	pc	INCLUSIVE OF INSTALLATION Fetal Doppler Specification: Display: Signal quality Display Heart Rate Range: approx. 50~240 bpm \pm 2% Ultrasound Frequency: approx. 2MHz / 3MHz Audio Output: at least 1.2W (Built in Loudspeaker), at least 3.5mm Phone jack Auto Shut Off: Sound: 1 min, Power: 5min Power: 1.5 V Battery X 2 (Type : AA)	
71	18	pc	Foot stool Specification: Strong tubular frame work. Sides of the top are covered with aluminum angles Non-slip rubber tips on stool legs. Non-slip corrugated rubber step tread. Dimension: approx. L: 60cm x W: 35cm x H: 40cm	
72	1	unit	Forcep jar/container without cover, stainless, size: 55mm x 180mm Fully motorized Dental Chair and unit Specification: Chair and Unit: Motorized Adjustment: Electric-powered adjustments for chair positioning (height, backrest, leg rest, recline) Size (L x W x H): Approximately 1300mm x 700mm x 1200mm Chair Upholstery: High-quality, seamless, easy-to-clean material (e.g., PU leather or vinyl)	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
			<p>Armrests: Adjustable, ergonomic armrests on both sides. Footrest: Adjustable, with sturdy support. Headrest: Adjustable, with comfortable padding and support. Base: Stable, heavy-duty base with locking mechanism for safety. Reclining Angle: Up to 180° for full comfort during procedures.</p> <p>Unit Configuration: Foot Control: For easy operation of chair movements (height, recline, leg rest). Backrest: Adjustable backrest with electric motor for precise positioning. Operating Light: LED light with adjustable intensity for optimal illumination during procedures. Cuspidor: Rotatable with standard water spray and suction system. Pressure Stabilizing Valve: Standard valve for regulating air pressure in handpieces.</p> <p>Power Supply: Voltage: at least 110V or 220V</p> <p>Accessories: Handpieces: 1 x High-speed handpiece 1 x Low-speed handpiece X-ray Viewing: 1 x Periapical x-ray viewer Film Viewer: 1 x 24V Film viewer for patient x-rays Storage: At least 6 shelves for storing dental instruments and materials</p> <p>INCLUSIVE OF INSTALLATION/COMMISSIONING</p>	
73	16	pc	<p>Glucometer with Blood Glucose Strips (50 strips/box) with normal and abnormal control</p>	
74	2	pc	<p>Graduated cylinder, glass, 500ml</p>	
75	4	set	<p>Gyne Set 10 Kelly forcep 7" 4 Allis forcep 6" 1 Tissue forcep 5½" 1 Thumb forcep 5½" 1 Debakey forcep 7¼" 1 Needle holder 6" 1 Needle holder 7" 1 Scalpel 5¼" 2 Bobcock forcep 6¼" 2 Mixer forcep 7¼"</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
76	21	unit	<p>5 Towel Clips 5¼"</p> <p>2 Army Navy 8¼"</p> <p>2 Richardson 9½", ¾ x 1 grip handle</p> <p>1 Malleable retractor 1x½x13"</p> <p>5 Ochsner curve 6 3/4"</p> <p>5 Ochsner straight 6 3/4"</p> <p>1 Long thumb 8"</p> <p>Heavy duty nebulizer</p> <p>Specification:</p> <p>Size: approx. 10.1" W x 10.5" H x 6.5" D</p> <p>Weight: at least 7-lbs</p> <p>Power requirements: approx. 1.3 amps, 115VAC 60Hz</p>	
77	2	set	Hegars dilator set of 8, stainless	
78	1	pc	<p>Hematocrit reader</p> <p>Specification:</p> <p>Material: Durable plastic, Non slip base, clear scale markings</p> <p>Dimensions: Approx. 5" (L) x 3" (W) x 1" (H), compact and easy to handle.</p>	
79	3	pc	Hemocytometer/Counting chamber, approx. 74mmx33mmx5mm	
80	2	set	<p>Herniorrhaphy Set</p> <p>10 Kelly forcep 6"</p> <p>4 Allis forcep 6"</p> <p>1 Tissue forcep 5½"</p> <p>1 Thumb forcep 5½"</p> <p>1 Debakey forcep 6"</p> <p>1 Needle holder 6"</p> <p>1 Needle holder 7"</p> <p>1 Scalpel 5¼"</p> <p>2 Bobcock 6¼"</p> <p>5 Towel Clips 5½"</p> <p>2 Army Navy 8¼"</p>	
81	100	unit	<p>Hospital Bed with Bed Mattress</p> <p>Specification:</p> <p>3 crank system</p> <p>double protection screw system</p> <p>polypropylene head and foot board</p> <p>aluminum side rail</p> <p>with silent castor system</p> <p>Dimensions: approximately 215 cm x 95 cm</p> <p>Mattress made of high quality polyurethane foam</p> <p>with class A leatherette cover</p> <p>color: blue</p> <p>Dimensions: approx. 36" (W) x 80" (L)</p>	

**Technical Specification
PR No. 25 - 0977**

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
82	10	unit	<p>Hydraulic Emergency Room Stretcher Bed</p> <p>Specification:</p> <p>Frame material: Sturdy stainless steel or high-quality carbon steel</p> <p>Top Coat: Electrostatic powder coating</p> <p>Full Length: at least 210cm, Full Width: at least 85cm</p> <p>Bed base height range: approx. 63cm-95cm</p> <p>Back section panel: approx. 0~75°</p> <p>Trendelenburg approx. 0~15°, Anti-Trendelenburg approx. 0~15°</p> <p>Diameter of Centrally Control Silent and Dustproof Castor: at least 20cm</p> <p>With Two Separate Hydraulic Pump</p> <p>With I.V.Pole</p>	
83	12	set	<p>Hysterectomy Set</p> <p>10 Kelly forcep 7"</p> <p>4 Allis forcep 6"</p> <p>1 Tissue forcep 5½"</p> <p>1 Thumb forcep 5½"</p> <p>1 Debakey forcep 7¼"</p> <p>1 Needle holder 7"</p> <p>1 Needle holder 10"</p> <p>1 Scalpel 5¼"</p> <p>1 Bobcock forcep 6¼"</p> <p>1 Mixer forcep`</p>	
84	7	unit	<p>Incubator/Infant Warmer</p> <p>Specification:</p> <p>Temperature Control Modes:</p> <p>Modes of Temperature Control: Skin Temperature Mode and Air Temperature Mode.</p> <p>Air Mode Control Temperature Range: Adjustable range from 20°C to 37°C (68°F to 98.6°F), with an override range up to 39°C (102.2°F).</p> <p>Skin Mode Control Temperature Range: Adjustable range from 34°C to 37°C (93.2°F to 98.6°F), with an override range up to 38°C (100.4°F).</p> <p>Physical Attributes:</p> <p>Dimensions: at least 140(H)x120.6(W)x75(D)cm</p> <p>Weight: at least 95.5 kg</p> <p>Hood Specifications:</p> <p>Access Panel: front and rear access panel</p> <p>Ports: at least 4 access ports and 2 iris ports</p> <p>Access panel opening height: at least 28.0 cm</p> <p>Mattress tray size: approx. 40.6 x 81 cm</p> <p>Mattress to hood height: at least 41.2 cm</p> <p>Mattress size: approx. 38.1 cm x 73.66 cm</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
85	14	unit	<p>Infusion Pump</p> <p>Specification:</p> <p>Infusion Rate: approx. 1-1200ml/h</p> <p>Flow Rate: approx. 100-1200ml/h</p> <p>Air Bubble size: at least \leq ul</p>	
86	8	unit	<p>Injection Pump</p> <p>Specification:</p> <p>Compatible between 10ml-60ml capacity, from various brands</p> <p>Total Injection Volume: at least 0.1ml</p> <p>Freely stackable</p>	
87	10	pc	Inoculating loops, sterile, disposable	
88	10	pc	Inoculating needles, sterile, disposable	
89	10	unit	<p>Instrument trolley/Back table</p> <p>Specification:</p> <p>Dimensions: approx. 24"x60"x34"</p> <p>Heavy duty with solid welded construction</p> <p>Stainless steel</p> <p>At least 3" ball bearing casters</p>	
90	4	pc	Intestinal clamp, curve 7", stainless	
91	4	pc	Intestinal clamp, straight 7", stainless	
92	4	pc	Intestinal clamp, curve 5", stainless	
93	4	pc	Intestinal clamp, straight 5", stainless	
94	5	pc	Iris scissor, curve 4", stainless	
95	5	pc	Iris scissor, straight 4", stainless	
96	100	pc	<p>IV stand with wheels, 4 hooks</p> <p>Specification:</p> <p>Easy rolling wheel ball bearing nylon casters</p> <p>Height range approx. 52.5" – 94"</p> <p>Twist lock for easy height adjustment</p>	
97	26	pc	Kelly hemostatic forceps, curve 7", stainless	
98	34	pc	Kelly hemostatic forceps, straight 7", stainless	
99	3	pc	Kelly pad, rubber, standard size	
100	40	pc	Kidney basing 12", stainless	
101	9	unit	<p>Kidney table, stainless steel</p> <p>Specification:</p> <p>Shape: semicircular</p> <p>Material: stainless steel</p> <p>Shelves: at least 2 shelves</p> <p>Dimensions: at least 60mm upraised back edge</p>	
102	2	pc	Kielland obstetrical forceps 16 1/4", stainless	
103	1	set	<p>LARYNGOSCOPE W/ DIFFERENT SIZES OF BLADES(NEONATE)</p> <p>4 blades</p> <p>Blade size: 0,00</p> <p>Stainless steel</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
104	5	set	LARYNGOSCOPE W/ DIFFERENT SIZES OF BLADES(PEDIA) 4 Blades Blade sizes 0,1, 2, 3 Stainless steel	
105	6	set	LARYNGOSCOPE WITH DIFFERENT SIZES OF BLADES (ADULT) 4 Blades Blade sizes 1,2, 3, 4 Stainless steel	
106	2	pair	Lead gloves Standard size at least 0.5 mmPB	
107	10	pc	Lead gown L 14.5 x W 21 x H 36 at least 0.35mm PB	
108	1	unit	Light cure machine Specifications: Input Voltage: approx. 110- 220V Wavelength: approx. 420-480nm	
109	3	pc	Lona pad/ Rubber sheet, 1 yard	
110	10	pc	Long thumb forcep 8", stainless	
111	2	pc	Luxator L3S, straight blade, stainless steel, at least 3mm	
112	2	pc	Luxator L5S, curved blade, stainless steel, at least 5mm	
113	3	set	MAJOR SET 10 Kelly forcep 7" 4 Allis forcep 6" 1 Tissue forcep 5½" 1 Thumb forcep 5½" 1 Debakey forcep 7¼" 1 Needle holder 6" 1 Needle holder 7" 1 Scalpel 5¼" 2 Bobcock forcep 6¼" 2 Mixer forcep 7¼" 5 Towel Clips 5¼" 2 Army Navy 8¼" 2 Richardson 9½", ¾ x 1 grip handle 1 Malleable retractor 1x½x13" 1 Intestinal clamp set 1 Long thumb 8"	
114	10	pc	Malleable retractor 1 1/2 x 13", stainless	
115	10	pc	Malleable retractor 75mmx13", stainless	
116	2	pc	Mallet, Head Diameter: approx. 29mm x 185mm, stainless steel	
117	20	pc	Mayo scissor, curve 6 3/4", stainless	
118	45	pc	Mayo scissor, straight 6 3/4", stainless	
119	14	pc	Mayo stand with tray, 4-wheels, stainless, with knob control	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
120	4	unit	<p>Mechanical Ventilator with Continuous Positive Airway Pressure (CPAP), Adult</p> <p>Specification:</p> <p>SCREEN DISPLAY: at least 12.1" color active matrix touchscreen, at least 1280x800 pixels</p> <p>TIDAL VOLUME: approx. 20-2000 mL</p> <p>FREQUENCY: approx. 1-100 BPM</p> <p>Tslope (Time of Pressure Rising): approx. 0.00-2.00 s</p> <p>Thigh: approx. 0.10-30 s</p> <p>Tlow: approx. 0.20-30.00 s</p> <p>O2 Therapy</p> <p>O2%: approx.21 to 100%</p> <p>Flow: Adult: approx. 2 to 80 L/min</p> <p>Maximum leakage compensation flow: Adult: 65 L/min</p> <p>Built in Nebulizer</p> <p>Pulmosight: Graphically displays resistance, compliance and breathing state, prompts abnormal compliance</p>	
121	30	pc	Metzenbaum scissor, curve 7", stainless	
122	30	pc	Metzenbaum scissor, straight 7", stainless	
123	26	set	<p>MINOR SET</p> <p>10 Mosquito curved forcep 5"</p> <p>2 Allis forcep 6"</p> <p>1 Thumb forcep 5½"</p> <p>1 Tissue forcep 5½"</p> <p>1 Needle holder 6"</p> <p>1 Scalpel 5¼"</p> <p>2 Skin retractor 6¼"</p> <p>2 Army navy 8¼"</p>	
124	10	pc	Mixer forcep, 7", stainless	
125	10	pc	Mixer forcep, 10", stainless	
126	2	pc	Molt Periosteal Elevator # 9, curved blade, 7.5mm wide	
127	1	pc	Mortar and pestle, glass, at least 90 x 50mm	
128	36	pc	Mosquito forceps, curve, 5", stainless	
129	48	pc	Mosquito forceps, straight, 5", stainless	
130	3	set	<p>Multigravida set</p> <p>2 Kelly forceps 7"</p> <p>1 Tissue forcep 5½"</p> <p>1 Thumb forcep 5½"</p> <p>1 Needle holder 6"</p> <p>7 Bandage scissors 7¼"</p>	
131	5	pc	Needle Holder 6", heavy	
132	5	pc	Needle Holder 6", light	
133	5	pc	Needle Holder 8", heavy	
134	5	pc	Needle Holder 8", light	
135	2	pc	<p>Negatoscope, double viewer, stainless steel,</p> <p>Size: at least 52"x28"x19"</p>	
136	8	pc	<p>Negatoscope, single viewer, stainless steel,</p> <p>Size: at least 17"x21"x2.5" deep</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
137	1	unit	<p>Neonatal Ventilator with Continuous Positive Airway Pressure (CPAP)</p> <p>Specification:</p> <p>SCREEN DISPLAY: at least 12.1" color active matrix touchscreen, at least 1280x800 pixels</p> <p>TIDAL VOLUME: approx. 20-2000 mL</p> <p>FREQUENCY: approx. 1-100 BPM</p> <p>Tslope (Time of Pressure Rising): approx. 0.00-2.00 s</p> <p>Thigh: approx. 0.10-30 s</p> <p>Tlow: approx. 0.20-30.00 s</p> <p>O2 Therapy</p> <p>O2%: approx.21 to 100%</p> <p>Flow: Neonate/Pediatric: approx. 2 to 80 L/min</p> <p>Maximum leakage compensation flow: Pediatric: 45 L/min</p>	
138	4	pc	<p>Built in Nebulizer</p> <p>Neonate, Weighing Scale</p> <p>Specification:</p> <p>Material: ABS/Metal</p> <p>Dimensions Unit: approx. 54.5×34.5×13.5 cm</p> <p>Shipping Weight: at least 4 kg</p>	
139	16	pc	Neurologic hammer with pin & brush, stainless, standard size	
140	10	pc	Ochsner forceps curve 6 3/4", stainless	
141	10	pc	Ochsner forceps straight 6 3/4", stainless	
142	1	unit	<p>Operating Table for Minor Surgery with Stirrups</p> <p>Specification:</p> <p>Material: stainless steel</p> <p>Sections: detachable head, leg & pelvic section</p> <p>Kidney bridge: built in kidney bridge</p> <p>Pillar Design: center pillar</p> <p>Length of table: at least 1900mm</p> <p>Width of table top: at least 520mm</p> <p>Min height (w/out mattress): 750mm</p> <p>Max height (w/out mattress): 1000mm</p> <p>Floor locking system</p> <p>INCLUSIVE OF INSTALLATION</p>	
143	1	unit	<p>Operating Table Gynecological with Stirrups</p> <p>Specification:</p> <p>Material: Base shell and stand column made of stainless steel.</p> <p>Height Adjustments:</p> <p>Table Height Range: approx. 650mm~895mm</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
144	1	unit	<p>Dimensions: Length and width: approx. 1780mm x 600mm Back section: at least 600mm x 800mm Seat section: at least 600mm x 400mm Leg section: at least 540mm x 500mm</p> <p>Trendelenburg / Reverse: at least 10°/25°</p> <p>Back Section up/down: at least 75°/6° INCLUSIVE OF INSTALLATION</p> <p>Operating Table Multipurpose Electro-Hydraulic Specification:</p> <p>Longitudinal Slide: at least 320 mm for free access to C-arm</p> <p>Design: Multi-layer decompression pad with waterproof seamless design Corded Remote Control</p> <p>Interchangeable Head plates and Leg plates Whole table top should be made of radiolucent material With two standard built-in battery, can support approximately 60 to 80 surgeries Back-up control system on the column Optical indicator for low battery reminder Power Supply: approx.100-240 VAC, 50/60 Hz</p> <p>Dimensions: Table length: at least 2060 mm Height without pad: approx. 498-1,000 mm Table Width without side rail: at least 520 mm Table Width with side rail: at least 590 mm Trendelenburg/Reverse Trendelenburg: approx. 25° Left lateral tilt (left/right): approx. 20° Head plate (up/down): +45°/ -90° Back plate (up/down): +80°/-40° Leg plates (up/down): +20°/-90° Flex/Reflex: 220°/110° Max. lifting capacity: 380 kg. Max. safe working load 250 kg.</p> <p>Accessories: head plate left leg plate right leg plate memory foam pad hand control with shift function light Anesthetic frame w/clamp heavy arm board w/clamp (MFP)</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
145	1	unit	<p>Advanced body strap INCLUSIVE OF INSTALLATION</p> <p>Operating Table with Orthopedic Surgery Capability Specification:</p> <p>Longitudinal Slide: at least 320 mm for free access to C-arm Maximum lifting capacity up to 380 kg. with highest safety and stability Maximum safe working load 250 kg.</p> <p>Multi-layer decompression pad with waterproof seamless design Corded Remote Control Interchangeable Head plates and Leg plates Whole table top should be made of radiolucent material</p> <p>Battery and Power:</p> <p>With two standard built-in battery, can support approximately 60 to 80 surgeries Power Supply: approx. 100-240 VAC, 50/60 Hz</p> <p>Dimensions: Table length: at least 2060 mm Height without pad: approx. 498-1,000 mm Table Width without side rail: at least 520 mm Table Width with side rail: at least 590 mm</p> <p>Accessories: head plate left leg plate right leg plate memory foam pad hand control with shift function light Anesthetic frame w/clamp heavy arm board w/clamp (MFP) Advanced body strap</p> <p>ORTHOPEDIC ACCESSORIES: LEG SURGERY: Orthopedic Extension Table Tibia and fibula Assy Lateral Femur Assy Transfer leg plate Unilateral lithotomy position Assy Leg Support L-shaped traction extension Arthroscopic leg holder Forearm extension device Side rail extension (L-shaped)</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
146	1	unit	<p>Upper arm extension device Arm operating table</p> <p>SPINE SURGERY: Bow frame for spine operation Bow frame cart Bow frame disposable cover (6 sets) Pad set for bow frame, pair Foam head block stand of bow frame</p> <p>HEAD SURGERY: Universal adaptor for head position Aluminum headrest Adult horseshoe headrest w/ extension Pedia horseshoe headrest w/ extension Stainless steel skull pin, adult (3 pcs) Stainless steel skull pin, pedia (3 pcs)</p> <p>INCLUSIVE OF INSTALLATION Operating Table with Sitrups Secification:</p> <p>Longitudinally movement: at least 350mm, which can fulfill the needs of C type arm whole body examination; Table width: at least 540mm; Leg board extension: 180°,folded 90°and detached; Accessories, guide rail, base shell and column shell all are made of stainless steel Table is equipped Auto-Charge Device, It can keep the table being used when emergency power shut off.</p>	
147	1	unit	<p>INCLUSIVE OF INSTALLATION OR LIGHT CEILING 1 ARM >160,000 lux Specification: Intensity: at least 160,000 Lux ±10% Size of Light Field: at least 12-30 cm Colour Temperature: approx. 4000-5000 K Colour Reduction Index: at least 93 RA LED Life: >50000 Hrs. Number of LED: at least 126 Diameter of Light: at least 500mm Brightness Control: Capacitive Touch Panel Power Supply: 220V/50Hz AC Colour of LED: White & Yellow (Double Colour) INCLUSIVE OF INSTALLATION</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
148	1	unit	<p>OR LIGHT CEILING 3 ARM >160,000 lux</p> <p>Specification:</p> <p>Variable Color Temperature</p> <p>M-Field: uses Bionic "compound eye structure" lens design to achieve a large light field of 60cm</p> <p>Touch Screen Control</p> <p>Infection Control Design: Screwless design, Laminar flow design, Antibacterial coating surface</p> <p>Max. Illuminance (Ec) (1m): at least 160,000 lux</p> <p>Light field diameter (1m): approx. 140-300mm</p> <p>Depth of illumination (20%): at least 1,400 mm</p> <p>Depth of illumination (60%): at least 800 mm</p> <p>Color Temperature: approx. 3,500-5,100K</p> <p>Power supply of all light sources: at least 30w</p>	
149	2	set	<p>INCLUSIVE OF INSTALLATION</p> <p>ORTHOPEDIC SET</p> <p>6 Kelly forcep 7"</p> <p>2 Bone holding forcep</p> <p>1 Curette</p> <p>1 Pliers</p> <p>1 Bone Rongeur</p> <p>1 Bone Cutters</p> <p>1 Chissel</p> <p>1 Gigli saw/wire</p> <p>2 Hohmann bone elevator</p> <p>1 Orthopedic bone curette</p> <p>1 Periosteal elevators</p> <p>1 Scalpel 5¼"</p> <p>1 Thumb forcep 5½"</p> <p>1 Tissue forcep 5½"</p> <p>5 Towel clips 5¼"</p> <p>1 Weitlaner retractor</p>	
150	100	pc	<p>Over bed table</p> <p>Specification:</p> <p>Frame material: Epoxy coated steel frame</p> <p>Height adjustment: The height can be adjusted by gas spring approx. 680 to 980mm.</p> <p>Base and Casters: diameter of at least 50mm castors, two of them with brakes</p> <p>Weight capacity: at least 10Kg</p>	
151	11	pc	Ovum forceps 9 1/2", stainless	
152	59	pc	<p>Oxygen regulator with flow meter (Tank and Wall Compatible)</p> <p>Specification:</p> <p>Inlet Pressure (Tank Mode): approx. 2,000 – 3,000 PSI</p> <p>Inlet Pressure (Wall Mode): approx. 50 – 60 PSI (hospital standard)</p> <p>Adjustable Outlet Pressure: approx. 0 – 100 PSI</p> <p>Flow Rate: approx. 0 – 15 LPM</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
153	29	pc	<p>Outlet Connection: Barbed hose fitting or Diameter Index Safety System (DISS) male fitting</p> <p>Material & Build: Body Material: Brass or chrome-plated brass Gauge Material: Impact-resistant polycarbonate lens for high visibility</p> <p>Oxygen tank holder, with trolley for 50 lbs, stainless</p>	
154	7	unit	<p>Patient Transport Stretcher</p> <p>Specification:</p> <p>Functionality: 2-function Design: backrest by qualified airspring, height adjustment by solid crank system</p> <p>Accessories Details: Metal dust protection for crank system, more durable than plastic cover. Cover all metal screws with rubber cap, protect screws from possible corrosion 4 hooks for IV pole Back lift can be up to 80 degrees Four double side casters of 6 inch diameter, simultaneously lockable by the step pedal at both sides. Stainless steel handle, more durable</p> <p>Adjustable vertical oxygen bottle, load capacity: at least 15KG</p> <p>Parameter Size: 1900*640*550-850mm Back-rest lifting: approx. 0-75° Loading bearing: at least 220kg</p>	
155	2	pc	Peak flow meter, adult, plastic tube, manual, 60-800L/mn	
156	2	pc	Peak flow meter, pedia, plastic tube, manual, 50-400L/mn	
157	1	unit	<p>Pediatric Suction Machine Heavy Duty</p> <p>Specification:</p> <p>Negative pressure: approx. 0.02-0.09MPa Noise: at least ≤65dB (A) Power: at least ≤180VA Pumping rate: at least ≥20L/min Reservoir capacity: at least 2500mL/pc, 2pieces Power Voltage: approx. AC220V±22V, 50Hz±1Hz</p>	
158	8	pc	Penlight (LED Light)	
159	4	pc	Pennington 6", stainless	
160	4	pc	Pennington 7", stainless	
161	3	pc	Petri dish with cover, glass, at least 90mm	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
162	1	unit	Pipette shaker Specification: Timer options: Adjustable timer or continuous operation Pipette holder: made of stainless steel spring clips Capacity: at least 6 pipettes Speed: at least 2500 rpm	
163	80	pc	Pipette Tips for 10-100ul and 100-1000ul	
164	4	pc	Pipettor Micropipette, 100-1000ul	
165	4	pc	Pipettor Micropipette, 10-100ul	
166	9	unit	Portable OR Light Specification: Type: mobile/ portable/ floor/ standing type surgical light (operating room light) Illumination: approx.120,000-160,000 lux Battery: at least 3 hours of use Color temperature: approx. 3500-5500K Head diameter: at least 500mm (19.7" inches) Whole weight: at least 40kg	
167	1	unit	Portable Dental XRAY with movable film Specification: Tube voltage: at least 60KV Tube current: at least 2mA Exposure time: approx. 0.1-2.0s Frequency: at least 30KHz Rate power: at least 120VA Distance from skin to cone: at least 118mm Tube focus: approx. 0.8mm x 0.8mm Battery: at least 2200mAh	
168	3	set	Primigravida set Inclusion: 1/4"1 pc. Operating Scissors S/B 5 1/2" straight 1 pc. Braun Episiotomy Scissors 8 1/2" Angled to side 1 pc. Umbilical Scissors 4" 2 pcs. Thumb Forceps 5 1/2" Serrated 2 pcs. Tissue Forceps (1x2T) 5 1/2" 6 pcs. Allis Tissue Forceps (5x6T) 6" 6 pcs. Crile Forceps 5 1/2" straight 1 pc. Bozeman Uterine Dressing Forceps 10 1/4" 2 pcs. Schroeder Braun Uterine Tenaculum 10 1/4" straight (Double curved) 1 pc. Lowa Membrane Puncturing Forceps 10 1/4" (Double Curved) 1 pc. Simpson Obstetrical Forceps 4" 1 pc. McLean-Tucker-Luikart Obstetrical Forceps 15 1/2" 1 pc. Kielland Obstetrical Forceps 17 1/2" 1 pc. Piper Obstetrical Forceps 17 1/2" 1 pc. Bill Traction Handle (For use with any Obstetrical Forceps)	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
			1 pc. Martin Pelvimeter (Graduated in cm and inches)	
			1 pc. Mayo-Hegar Needle Holder 6"	
			1 pc. Mayo- Hegar Needle Holder 7"	
			1 pc. Sterilizing tray (Perforated)	
169	40	pc	Pulse oximeter fingertip, adult	
170	12	pc	Pulse oximeter fingertip, pedia	
171	4	pc	Pulse oximeter, fingertip, neonate	
172	6	pc	Randall forceps 9", stainless	
173	10	pc	RBC micropipette Specification: Material: glass Markings: Includes standard markings at 0.5, 1, and 11 for accurate measurements. Size: Standard size.	
174	10	pc	Russian forceps 8", stainless	
175	10	pc	Russian forceps 9", stainless	
176	9	pc	Scalpel handle, no. 3, stainless	
177	9	pc	Scalpel handle, no. 4, stainless	
178	3	unit	Scrub Sink Stainless, Single Specification: Material: Sink made of heavy-duty stainless steel sheet for durability and easy cleaning. Foot Pedal Operated: Features a foot pedal operated mechanism for hands-free operation Size: At least 22" x 24" x 35" Sink Depth: Sink depth of at least 15 inches INCLUSIVE OF INSTALLATION	
179	5	unit	Scrub Sink Stainless 2 faucet and foot pedal Specification: Material: Sink made of heavy-duty stainless steel sheet. Frame: Stand made of stainless steel tube, ensuring durability. Foot Pedal Operated: Foot pedal mechanism for hands-free operation. Faucets: Equipped with 2 faucets for dual use. Size: At least 22 x 24 x 35 inches. Sink Depth: Sink depth of at least 15 inches for comfortable use. INCLUSIVE OF INSTALLATION	
180	5	pc	Senn retractor, stainless, standard size	
181	10	pc	Serological pipette, 10mL, glass	
182	10	pc	Serological pipette, 1mL, glass	
183	10	pc	Serological pipette, 20mL, glass	
184	10	pc	Serological pipette, 5mL, glass	
185	9	pc	Sharp disposable container, yellow, at least 5L	
186	1	pc	Simpson braun obstetrical forceps, curve, 14 1/2", stainless	

**Technical Specification
PR No. 25 - 0977**

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
187	1	pc	Simpson braun obstetrical forceps, straight, 14 1/2", stainless	
188	22	pc	Soaking tray/pan, big, stainless	
189	20	pc	Soaking tray/pan, small, stainless	
190	21	unit	Sphygmomanometer with trolley, non-mercurial, adult cuff	
191	12	unit	Sphygmomanometer with trolley, non-mercurial, pedia cuff	
192	1	pc	Spirometer, at least 4000ml with one-way valve	
193	24	pc	Staining glass/Coplin jar, standard size	
194	1	pc	Staining rack	
			Specification:	
			Capacity: at least 60	
			Size: at least 1"x3"	
195	1	unit	Stationary Digital XRAY Machine	
			Specification:	
			GENERATOR	
			Maximum output power: at least 50 kW	
			kVp range: approx. 40 - 150 kVp	
			mA range: approx. least 10-630 mA	
			Exposure time range: approx. 2 - 2000 ms	
			mAs range: up to 630 mAs	
			AEC nominal irradiation shortest time: at least 2 ms	
			X-RAY TUBE	
			Tube type: high speed, rotating anode tube	
			Target angle: at least 12 degrees	
			Anode Rotation Speed: Min. 9,700 rpm (180 Hz)	
			Focal spot size: at least 0.6 / 1.2mm	
			Nominal anode power: at least 32 kW for small focal and 78 kW for large focal	
			Anode heat content: 210 kJ (Anode) and 900 kJ (tube assembly)	
			Cooling and protection: unit must have thermal switch and pressure switch.	
			Tube Weight: not more than 18kg	
			Static Patient Table	
			Type: fully flat top, static table	
			Tabletop size: approx. 85.7 x 235 cm (33.7 x 92.5 in)	
			Tabletop material: carbon-fiber composite	
			Weight limit: at least 200 kg patient centered on the table	
			Includes foot pedal with single click to release	
			Grid: at least 100 cm with focal range of 90-120cm; 70 lp/cm must support at least 17 x 17 inches flat panel detector	
			Manual Wall Stand	
			Type: Manual vertical movement	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
			<p>Floor to center of detector: at least of 50 cm Vertical travel range: at least 126 cm Motion safety: vertical electromagnetic braking system Grid: at least 130 cm with focal range of 90-190cm; 70 lp/cm Detector size: approx. 43 x 43 cm (17 x 17 in)</p> <p>Dimension: approx. 460 x 460 x 15 mm Pixel Matrix: approx. 4267 x 4267 pixels Pixel size: at least 100 µm Load Capacity: max 300 kg Battery operation must be at least 8 hours</p> <p>INCLUSIVE OF INSTALLATION & COMMISSIONING</p>	
196	33	pc	Stethoscope, adult	
197	19	pc	Stethoscope, pedia	
198	5	pc	Straight forceps 6", stainless	
199	5	pc	Straight forceps 8", stainless	
200	26	unit	<p>Suction Machine Heavy Duty Specification:</p> <p>Design: Plastic panel construction Control: Operated with both hand-switch and foot-switch for user convenience.</p> <p>Performance: Negative Pressure: Adjustable within the typical range of 0.02-0.09 MPa Noise Level: Designed for quieter operation with a noise level of ≤ 65 dB(A). Power Consumption: Energy-efficient with a typical power consumption of ≤ 180 VA. Pumping Rate: Offers a suction capacity of ≥ 20 L/min for effective performance.</p> <p>Capacity: Reservoir: Includes 2 reservoirs, each with a typical capacity of 2500 mL for continuous operation.</p> <p>Power Supply: Voltage: Standard operating voltage of AC 220V ± 22V, 50Hz ± 1Hz</p>	
201	4	unit	<p>Supply and Installation of Anesthesia Machine Specification:</p> <p>Weight: Approximately 135 kg</p> <p>Power and Voltage: Power Consumption: Typically consumes less than 95W during standard operation, and can go up to 400W during mechanical ventilation.</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
202	2	pc	<p>Mains Voltage: Supports a wide range of power supplies, typically operating between 100 to 240V AC, with 50/60Hz frequency.</p> <p>Internal Battery Backup: Includes a battery with a typical backup time of at least 45 minutes to ensure uninterrupted use.</p> <p>Gas Supply: Central Gas Supply: Supports gas supplies such as Oxygen (O₂), Air (AIR), and Nitrous Oxide (N₂O), with typical supply pressures between 2.7 to 6.9 kPa (39 psi).</p> <p>Gas Cylinder Options: Can accommodate 1 or 2 standing gas cylinders. Option to use 2 or 3 suspended gas cylinders with a pin-index connector. Option to add a holder for 1 additional standing gas cylinder.</p> <p>Ventilator Specifications: Respiratory Rate (RR): Adjustable to a range between 3 to 100 breaths per minute. Inspiratory Time (Ti): Adjustable from approximately 0.2 to 10 seconds. Inspiratory Flow Rate: Supports a minimum flow of 0.1 L/min, with the ability to handle up to 160 L/min for more demanding cases</p> <p>Surgical handpiece Specification: Speed: at least 2,000 rpm (rotations per minute) Torque: At least 50 ncm (Newton-centimeters) Material: Stainless steel</p>	
203	3	set	<p>Design and Ergonomics: Compact and Lightweight Non-slip Grip</p> <p>Suture removal set Inclusion: Littauer Stitch Scissors 4 1/2" Adson Dressing Forceps 4 3/4" Kelly Forceps Straight 5 1/2" PVP prep pad Alcohol prep pad Gauze pad 2x2</p>	
204	3	unit	<p>Table Top Autoclave Machine Specification: Structure: Stainless steel Sterilizer Volume: At least 24L capacity</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
			<p>Safety Features: Water Protection: Equipped with safe protection to prevent operation in the absence of water. Temperature & Pressure Protection: Includes auto-protection features for both over-temperature and over-pressure conditions to ensure safe operation</p> <p>Sterilization Process: Timer: Adjustable timing between 0-60 minutes to accommodate various sterilization requirements. Temperature Range: Capable of operating within a temperature range of 105°C to 134°C, suitable for a variety of medical and laboratory instruments</p>	
205	10	pc	Test tube brush (Medium)	
206	2	pc	Test tube holder, at least 30mm diameter	
207	6	pc	Test tube rack, stainless, at least 50 capacity	
208	100	pc	Test tubes 15ml	
209	100	pc	Test tubes 5ml	
210	2	pc	<p>Thermometer for ref temp, non-mercurial</p> <p>Specification:</p> <p>Dial Design: At least 2.5"</p> <p>Temperature Range: Accurately measures temperatures from 20°F to 80°F</p> <p>Material: Stainless steel</p> <p>Safety and Usability: Safe Temperature Zone Indicators: Equipped with clear temperature zone indicators to easily identify safe operating ranges</p>	
211	2	pc	<p>Thermometer for Rectal</p> <p>Specification:</p> <p>Measurement Range: Approx. 32°C to 43.9°C</p> <p>Response Time: Approximately 10 seconds for rectal measurements</p> <p>Display: Large, easy-to-read LCD screen.</p> <p>Probe: Flexible, waterproof tip for comfort and easy cleaning.</p> <p>Power Source: Battery-operated, with low battery indicator.</p> <p>Safety Features: Automatic shut-off to conserve battery life.</p>	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
212	46	pc	Thermometer, digital pencil-type Specification: Measurement Range: Approx. 32°C to 43.9°C Response Time: Approx. 60-120 seconds Display: Large, easy-to-read LCD screen Probe: Flexible, waterproof tip for comfort and easy cleaning Power Source: Battery-operated, with low battery indicator Safety Features: Automatic shut-off to conserve battery life	
213	4	pc	Thermometer for Oral Specification: Measurement Range: Approx. 32°C to 43.9°C Response Time: Approx. 10-60 seconds Display: Large, easy-to-read LCD screen Probe: Flexible, waterproof tip for comfort and easy cleaning Power Source: Battery-operated, with low battery indicator Safety Features: Automatic shut-off to conserve battery life	
214	18	pc	Thermometer, Thermoscan Specification: Measurement Range: Temperature Range: Approx. 34°C to 42.2°C Response Time: 2 seconds for a reading. Display: LCD Screen with easy-to-read digits Power Source: Battery operated, with low battery indicator Color-Coded Display: Displays readings in green, yellow, or red for easy fever interpretation Fever Alarm: Alert when a high temperature is detected. Memory Function: Recalls the last reading. Protective Case Included with lens filters	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
215	32	pc	Thumb forceps 5", stainless	
216	3	pc	Thyroid shield Specification: Material: Lead Equivalent: 0.50mm Pb (Lead) (standard) Construction: Lightweight and durable material Universal Size: Adjustable design fits a variety of neck sizes Thyroid Shielding: Specifically designed to protect the thyroid gland during radiographic or X-ray procedure	
217	3	set	Thyroidectomy Set 10 Mosquito forcep curved 5" 6 Hemostatic forcep 6" 4 Allis forcep 6" 1 Tissue forcep 5½" 1 Thumb forcep 5½" 1 DeBakey forcep 6" 1 Adson forcep 1 Tissue forcep smooth 4/3" 1 Needle holder 6" 1 Needle holder 7" 1 Scalpel 5 1/4" 2 Bobcock forcep 6 1/4" 2 Baby Mixer forcep 5" 5 Towel clips 5 1/2" 2 Army Navy 8 1/4" 1 Tracheal Retractor 1 Weitlaner Retractor	
218	47	pc	Tissue forceps 5", stainless	
219	80	pc	Towel clips 5", stainless	
220	3	set	Tracheostomy Set 4 Kelly forcep 5 1/2" 1 Tissue forcep 5 1/2" 1 Thumb forcep 5 1/2" 1 Metz Scissor 5 1/2" 3 Senn Retractor Prong sharp- 6 1/4" 2 Army navy 8 1/4" 1 Weitlaner 5" 1 Blade holder 5 1/4" 6 Tracheostomy 12"	
221	2	unit	Ultrasound bed, Hamilton type Specification: Design & Construction: Frame Type: Sturdy construction with a strong base and supportive structure. Top Surface: Upholstered three-section top Drawer & Compartment: Includes a storage drawer and compartment door for easy organization of accessories and equipment	

Technical Specification
PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
222	1	unit	<p>Stirrup: Equipped with stirrups for proper positioning during procedures.</p> <p>Sliding Foot Stool: Includes a sliding foot stool</p> <p>Dimensions: approx. 58(L) x 20(W) x 34(H) inches</p> <p>Upholstery: Material: High-quality, easy-to-clean upholstery designed for patient comfort.</p> <p>Functionality: Adjustable Sections: The three-section top allows for versatile positioning during ultrasound exams</p> <p>Ultrasonic scaler Specification: Dimensions: Approx.188mm (L) x 159mm (W) x 81mm (H)</p> <p>Output Power: Approx. 3W to 20W</p> <p>Compact Size: Designed with a compact form factor for easy handling and storage</p>	
223	4	pc	Umbilical scissor 4 1/4", stainless	
224	15	set	Vaginal speculum S,M,L, stainless	
225	12	pc	Vascular clamp 7", stainless	
226	11	set	Vascular dilator, stainless steel titanium, 21 cm long 0.5mm-4.0mm, 8pcs/set	
227	5	pc	Vein retractor, stainless, standard size	
228	1	unit	<p>Ventilator with Continuous Positive Airway Pressure (CPAP), Pedia</p> <p>Specification:</p> <p>Display: Screen Type: at least 12.1-inch color active matrix TFT touchscreen, at least 1280 x 800 pixels TIDAL VOLUME: approx. 20-2000 mL FREQUENCY: approx. 1-100 BPM Tslope (Time of Pressure Rising): approx. 0.00-2.00 s Thigh: approx. 0.10-30 s Tlow: approx. 0.20-30.00 s</p> <p>Oxygen Therapy: Oxygen Concentration (O2%): 21% to 100%. Oxygen Flow Rate: 2 to 80 L/min (adjustable for Adult/Pediatric). Maximum Leakage Compensation Flow: Adult: 65 L/min, Pediatric: 45 L/min.</p> <p>Nebulizer: Built-in Nebulizer</p>	

**Technical Specification
PR No. 25 - 0977**

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
229	1	unit	<p>White Blood Cell (WBC) differential counter</p> <p>Specification:</p> <p>Design: Manual or digital counters with multiple keys corresponding to different white blood cell types</p> <p>Key Configuration: Approx. 5-key or 8-key models</p> <p>Material: Durable construction, often with an ABS (Acrylonitrile Butadiene Styrene) outer shell</p> <p>Display: Mechanical counters with dials or digital counters with LCD screens to display counts.</p>	
230	10	pc	<p>White Blood Cell (WBC) Micropipette</p> <p>Specification:</p> <p>Material:</p> <p>Body: Glass</p> <p>Design:</p> <p>Bulb: White bead inside the bulb for visual identification and fluid measurement.</p> <p>Markings: Clearly visible scale markings at 0.5, 1, and 11 for accurate fluid measurements.</p> <p>Standard Size: Suitable for use in typical laboratory settings; standard size for WBC counting applications</p>	
231	12	pc	<p>Weitlaner sharp retractor 3x4 prongs, stainless</p>	
232	6	pc	<p>Weitlaner blunt retractor, 2x3 prongs, stainless</p>	
233	30	pc	<p>Wheel Chair</p> <p>Specification:</p> <p>Frame:</p> <p>Material: Stainless steel frame</p> <p>Construction: Frame constructed with 20-24 millimeter diameter tube, providing stability and support.</p> <p>Seat and Back:</p> <p>Material: Upholstered seat and back made of leatherette material</p> <p>Seat Rest: Reinforced seat rest with an integrated chart holder pocket for convenience during use.</p> <p>Footrest:</p> <p>Type: Folding aluminum footrest</p> <p>IV Pole:</p> <p>Material: Stainless steel IV pole with 2 hooks, adjustable for height to accommodate different patient needs.</p>	

Technical Specification PR No. 25 - 0977

Item No.	Quantity	Unit of Issue	Specification	Statement of Compliance*
			Rear Wheels: Non-magnetic rear mag wheels with solid tires for increased safety and durability, especially in hospital environments. Wheel Type: Designed to provide smooth mobility over different surfaces.	

*Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the the equipment offered. Statement of "Comply" or "Not Comply" must be supported by evidence in a Bidder's Bid and cross-referenced to that evidence. Evidence shall be in the the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or Supplier liable for prosecution subject to the provisions of ITB Clause 3.1 (a) (ii) and/or GCC Clause 2.1 (a) (ii).

I hereby certify to comply with the Technical Specifications of all the items to be Bid.

Name of Company/Bidder

Signature Over Printed Name of Representative

Date

SECTION VIII.
CHECKLIST OF
TECHNICAL &
FINANCIAL DOCUMENTS



Republic of the Philippines
 PROVINCE OF PAMPANGA
 City of San Fernando

BIDS AND AWARDS COMMITTEE
I.G.-BAC
BIDDING REQUIREMENTS FOR THE PROCUREMENT
OF GOOD/EQUIPMENTS

PROJECT

BIDDER

Checklist of BID Requirements

I. TECHNICAL COMPONENT ENVELOPE

Class "A" Documents

Legal Documents

□ □ □ □ □ □ □ □

(a.) Valid PhilGEPS Registration Certificate (Platinum Membership)(all pages);
in accordance with Section 8.5.2 of the IRR

Technical Documents

□ □ □ □ □ □ □ □

(b.) Statement of the prospective bidder of all its ongoing government and private contracts, including contracts awarded but not yet started, if any, whether similar or not similar in nature and complexity to the contract to be bid; **and**

□ □ □ □ □ □ □ □

(c.) Statement of bidder's Single Largest Completed Contract (SLCC) similar to the contract to be bid, except under conditions provided for in Sections 23.4.1.3 and 23.4.2.4 of the 2016 revised IRR of RA No. 9184, within the relevant period as provided in the Bidding Documents; **and**

□ □ □ □ □ □ □ □

(d.) Original Copy of Bid Security. If in the form of a Surety Bond, submit also a certification issued by the Insurance Commission;
or

□ □ □ □ □ □ □ □

Original Copy of Notarized Bid Securing Declaration; **and**
 (e.) Conformity with the Technical Specifications, which include production/delivery schedule; manpower requirements; aftersales; **and**

□ □ □ □ □ □ □ □

(f.) Original duly signed Omnibus Sworn Statement (OSS); **and** if applicable, Original Notarized Secretary's Certificate in case of a corporation, partnership, or cooperative; or Original Special Power of Attorney of all members of the joint venture giving full power and authority to its officer to sign the OSS and do acts to represent the Bidder.

Financial Documents

□ □ □ □ □ □ □ □

(g.) The prospective bidder's computation of Net Financial Contracting Capacity (NFCC);
or
 A committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation

Class "B" Documents

□ □ □ □ □ □ □ □

(h.) If applicable, a duly signed joint venture agreement (JVA) in case the joint venture is already in existence;
or
 duly notarized statements from all the potential joint venture partners stating
 Duly notarized statements from all the potential joint venture partners stating
 that the bid is successful.

Other documentary requirements under RA No. 9184 (as applicable)

--	--	--	--	--	--	--	--

(i.) [For foreign bidders claiming by reason of their country's extension of reciprocal rights to Filipinos] Certification from the relevant government office of their country stating that Filipinos are allowed to participate in government procurement activities for the same item or product.

--	--	--	--	--	--	--	--

(j.) Certification from the DTI if the Bidder claims preference as a Domestic Bidder or Domestic Entity.

II. FINANCIAL COMPONENT ENVELOPE

--	--	--	--	--	--	--	--

(a.) Original of duly signed and accomplished Financial Bid Form; **and**

--	--	--	--	--	--	--	--

(b.) Original of duly signed and accomplished Price Schedule(s)

NOTE : Any missing document/s in the above -mentioned checklist is a ground for outright declaration of "Non-Compliance" Suppliers and Contractors who will participate in the public bidding of the Province are required to submit One (1) Copy of the first & second components of their technical documents duly signed by the authorized signatory.

[] Complying

[] Non-Complying

SECTION IX.
OTHER BIDDING
DOCUMENTS

Bid Form for the Procurement of Goods

[shall be submitted with the Bid]

BID FORM

Date : _____
Project Identification No. : _____

To: [name and address of Procuring Entity]

Having examined the Philippine Bidding Documents (PBDs) including the Supplemental or Bid Bulletin Numbers *[insert numbers]*, the receipt of which is hereby duly acknowledged, we, the undersigned, offer to *[supply/deliver/perform]* *[description of the Goods]* in conformity with the said PBDs for the sum of *[total Bid amount in words and figures]* or the total calculated bid price, as evaluated and corrected for computational errors, and other bid modifications in accordance with the Price Schedules attached herewith and made part of this Bid. The total bid price includes the cost of all taxes, such as, but not limited to: *[specify the applicable taxes, e.g. (i) value added tax (VAT), (ii) income tax, (iii) local taxes, and (iv) other fiscal levies and duties]*, which are itemized herein or in the Price Schedules,

If our Bid is accepted, we undertake:

- a. to deliver the goods in accordance with the delivery schedule specified in the Schedule of Requirements of the Philippine Bidding Documents (PBDs);
- b. to provide a performance security in the form, amounts, and within the times prescribed in the PBDs;
- c. to abide by the Bid Validity Period specified in the PBDs and it shall remain binding upon us at any time before the expiration of that period.

[Insert this paragraph if Foreign-Assisted Project with the Development Partner:

Commissions or gratuities, if any, paid or to be paid by us to agents relating to this Bid, and to contract execution if we are awarded the contract, are listed below:

Name and address of agent	Amount and Purpose of Commission or gratuity
---------------------------	--

_____	_____
_____	_____
_____	_____

(if none, state "None")]

Until a formal Contract is prepared and executed, this Bid, together with your written acceptance thereof and your Notice of Award, shall be binding upon us.

We understand that you are not bound to accept the Lowest Calculated Bid or any Bid you may receive.

We certify/confirm that we comply with the eligibility requirements pursuant to the PBDs.

The undersigned is authorized to submit the bid on behalf of *[name of the bidder]* as evidenced by the attached *[state the written authority]*.

We acknowledge that failure to sign each and every page of this Bid Form, including the attached Schedule of Prices, shall be a ground for the rejection of our bid.

Name: _____

Legal capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____

Date: _____

**** Mandatory Provision pursuant to Government Procurement Policy Board (GPPB) Resolution No. 16-2020 dated September 16, 2020.***

Price Schedule for Goods Offered from Abroad
[shall be submitted with the Bid if bidder is offering goods from Abroad]

For Goods Offered from Abroad

Name of Bidder _____ Project ID No. _____ Page ___ of ___

1	2	3	4	5	6	7	8	9
Item	Description	Country of origin	Quantity	Unit price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 4 x 5)	Unit Price Delivered Duty Unpaid (DDU)	Unit price Delivered Duty Paid (DDP)	Total Price delivered DDP (col 4 x 8)

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____

*** Mandatory Provision pursuant to Government Procurement Policy Board (GPPB) Resolution No. 16-2020 dated September 16, 2020**

Price Schedule for Goods Offered from Within the Philippines

[shall be submitted with the Bid if bidder is offering goods from within the Philippines]

For Goods Offered from Within the Philippines

Name of Bidder _____ Project ID No. _____ Page __ of __

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if Contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 9) x (col 4)

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____

*** Mandatory Provision pursuant to Government Procurement Policy Board (GPPB) Resolution No. 16-2020 dated September 16, 2020.**

Bid Securing Declaration Form

[shall be submitted with the Bid if bidder opts to provide this form of bid security]

REPUBLIC OF THE PHILIPPINES) CITY OF
_____) S.S.

BID SECURING DECLARATION Project Identification No.: *[Insert number]*

To: *[Insert name and address of the Procuring Entity]*

I/We, the undersigned, declare that:

1. I/We understand that, according to your conditions, bids must be supported by a Bid Security, which may be in the form of a Bid Securing Declaration.
2. I/We accept that: (a) I/we will be automatically disqualified from bidding for any procurement contract with any procuring entity for a period of two (2) years upon receipt of your Blacklisting Order; and, (b) I/we will pay the applicable fine provided under Section 6 of the Guidelines on the Use of Bid Securing Declaration, within fifteen (15) days from receipt of the written demand by the procuring entity for the commission of acts resulting to the enforcement of the bid securing declaration under Sections 23.1(b), 34.2, 40.1 and 69.1, except 69.1(f), of the IRR of RA No. 9184; without prejudice to other legal action the government may undertake.
3. I/We understand that this Bid Securing Declaration shall cease to be valid on the following circumstances:
 - a. Upon expiration of the bid validity period, or any extension thereof pursuant to your request;
 - b. I am/we are declared ineligible or post-disqualified upon receipt of your notice to such effect, and (i) I/we failed to timely file a request for reconsideration or (ii) I/we filed a waiver to avail of said right; and
 - c. I am/we are declared the bidder with the Lowest Calculated Responsive Bid, and I/we have furnished the performance security and signed the Contract.

IN WITNESS WHEREOF, I/We have hereunto set my/our hand/s this ___ day of *[month]* *[year]* at *[place of execution]*.

[Insert NAME OF BIDDER OR ITS AUTHORIZED REPRESENTATIVE]

[Insert signatory's legal capacity] Affiant

[Jurat]

[Format shall be based on the latest Rules on Notarial Practice]

*** Mandatory Provision pursuant to Government Procurement Policy Board (GPPB) Resolution No. 16-2020 dated September 16, 2020.**

Contract Agreement Form for the Procurement of Goods (Revised)

[Not required to be submitted with the Bid, but it shall be submitted within ten (10) days after receiving the Notice of Award]

CONTRACT AGREEMENT

THIS AGREEMENT made the ____ day of _____ 20____ between [name of PROCURING ENTITY] of the Philippines (hereinafter called “the Entity”) of the one part and [name of Supplier] of [city and country of Supplier] (hereinafter called “the Supplier”) of the other part;

WHEREAS, the Entity invited Bids for certain goods and ancillary services, particularly [brief description of goods and services] and has accepted a Bid by the Supplier for the supply of those goods and services in the sum of *[contract price in words and figures in specified currency]* (hereinafter called “the Contract Price”).

NOW THIS AGREEMENT WITNESSETH AS FOLLOWS:

1. In this Agreement words and expressions shall have the same meanings as are respectively assigned to them in the Conditions of Contract referred to.

2. The following documents as required by the 2016 revised Implementing Rules and Regulations of Republic Act No. 9184 shall be deemed to form and be read and construed as integral part of this Agreement, viz.:

- i. Philippine Bidding Documents (PBDs);
 - i. Schedule of Requirements;
 - ii. Technical Specifications;
 - iii. General and Special Conditions of Contract; and
 - iv. Supplemental or Bid Bulletins, if any
- ii. Winning bidder’s bid, including the Eligibility requirements, Technical and Financial Proposals, and all other documents or statements submitted;

Bid form, including all the documents/statements contained in the Bidder’s bidding envelopes, as annexes, and all other documents submitted (*e.g.*, Bidder’s response to request for clarifications on the bid), including corrections to the bid, if any, resulting from the Procuring Entity’s bid evaluation;

iii. Performance Security;

iv. Notice of Award of Contract; and the Bidder’s conforme thereto; and

v. Other contract documents that may be required by existing laws and/or the Procuring Entity concerned in the PBDs. **Winning bidder agrees that additional contract documents or information prescribed by the GPPB that are subsequently required for submission after the contract execution, such as the Notice to Proceed, Variation Orders, and Warranty Security, shall likewise form part of the Contract.**

3. In consideration for the sum of *[total contract price in words and figures]* or such other sums as may be ascertained, *[Named of the bidder]* agrees to *[state the object of the contract]* in accordance with his/her/its Bid.

4. The *[Name of the procuring entity]* agrees to pay the above-mentioned sum in accordance with the terms of the Bidding.

IN WITNESS whereof the parties hereto have caused this Agreement to be executed in accordance with the laws of the Republic of the Philippines on the day and year first above written.

[Insert Name and Signature]

[Insert Name and Signature]

*[Insert Signatory's Legal Capacity]
for:*

*[Insert Signatory's Legal Capacity]
for:*

[Insert Procuring Entity]

[Insert Name of Supplier]

Acknowledgment

[Format shall be based on the latest Rules on Notarial Practice]

**** Mandatory Provision pursuant to Government Procurement Policy Board (GPPB) Resolution No. 16-2020 dated September 16, 2020.***

Omnibus Sworn Statement (Revised)

[shall be submitted with the Bid]

REPUBLIC OF THE PHILIPPINES)
CITY/MUNICIPALITY OF _____) S.S.

AFFIDAVIT

I, [Name of Affiant], of legal age, [Civil Status], [Nationality], and residing at [Address of Affiant], after having been duly sworn in accordance with law, do hereby depose and state that:

1. *[Select one, delete the other:]*

[If a sole proprietorship:] I am the sole proprietor or authorized representative of [Name of Bidder] with office address at [address of Bidder];

[If a partnership, corporation, cooperative, or joint venture:] I am the duly authorized and designated representative of [Name of Bidder] with office address at [address of Bidder];

2. *[Select one, delete the other:]*

[If a sole proprietorship:] As the owner and sole proprietor, or authorized representative of [Name of Bidder], I have full power and authority to do, execute and perform any and all acts necessary to participate, submit the bid, and to sign and execute the ensuing contract for [Name of the Project] of the [Name of the Procuring Entity], as shown in the attached duly notarized Special Power of Attorney;

[If a partnership, corporation, cooperative, or joint venture:] I am granted full power and authority to do, execute and perform any and all acts necessary to participate, submit the bid, and to sign and execute the ensuing contract for [Name of the Project] of the [Name of the Procuring Entity], as shown in the attached [state title of attached document showing proof of authorization (e.g., duly notarized Secretary's Certificate, Board/Partnership Resolution, or Special Power of Attorney, whichever is applicable)];

3. [Name of Bidder] is not "blacklisted" or barred from bidding by the Government of the Philippines or any of its agencies, offices, corporations, or Local Government Units, foreign government/foreign or international financing institution whose blacklisting rules have been recognized by the Government Procurement Policy Board, **by itself or by relation, membership, association, affiliation, or controlling interest with another blacklisted person or entity as defined and provided for in the Uniform Guidelines on Blacklisting;**
4. Each of the documents submitted in satisfaction of the bidding requirements is an authentic copy of the original, complete, and all statements and information provided therein are true and correct;
5. [Name of Bidder] is authorizing the Head of the Procuring Entity or its duly authorized representative(s) to verify all the documents submitted;

6. [Select one, delete the rest:]

[If a sole proprietorship:] The owner or sole proprietor is not related to the Head of the Procuring Entity, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

[If a partnership or cooperative:] None of the officers and members of [Name of Bidder] is related to the Head of the Procuring Entity, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

[If a corporation or joint venture:] None of the officers, directors, and controlling stockholders of [Name of Bidder] is related to the Head of the Procuring Entity, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

7. [Name of Bidder] complies with existing labor laws and standards; and
8. [Name of Bidder] is aware of and has undertaken the responsibilities as a Bidder in compliance with the Philippine Bidding Documents, which includes:
- a. Carefully examining all of the Bidding Documents;
 - b. Acknowledging all conditions, local or otherwise, affecting the implementation of the Contract;
 - c. Making an estimate of the facilities available and needed for the contract to be bid, if any; and
 - d. Inquiring or securing Supplemental/Bid Bulletin(s) issued for the [Name of the Project].
9. [Name of Bidder] did not give or pay directly or indirectly, any commission, amount, fee, or any form of consideration, pecuniary or otherwise, to any person or official, personnel or representative of the government in relation to any procurement project or activity.
10. **In case advance payment was made or given, failure to perform or deliver any of the obligations and undertakings in the contract shall be sufficient grounds to constitute criminal liability for Swindling (Estafa) or the commission of fraud with unfaithfulness or abuse of confidence through misappropriating or converting any payment received by a person or entity under an obligation involving the duty to deliver certain goods or services, to the prejudice of the public and the government of the Philippines pursuant to Article 315 of Act No. 3815 s. 1930, as amended, or the Revised Penal Code.**

IN WITNESS WHEREOF, I have hereunto set my hand this __ day of __, 20__ at _____, Philippines.

[Insert NAME OF BIDDER OR ITS AUTHORIZED REPRESENTATIVE]
[Insert signatory's legal capacity]
Affiant

* Mandatory Provision pursuant to Government Procurement Policy Board (GPPB) Resolution No. 16-2020 dated September 16, 2020.

Performance Securing Declaration (Revised)

[if used as an alternative performance security but it is not required to be submitted with the Bid, as it shall be submitted within ten (10) days after receiving the Notice of Award]

REPUBLIC OF THE PHILIPPINES)
CITY OF _____) S.S.

PERFORMANCE SECURING DECLARATION

Invitation to Bid: [Insert Reference Number indicated in the Bidding Documents] To:
[Insert name and address of the Procuring Entity]

I/We, the undersigned, declare that:

1. I/We understand that, according to your conditions, to guarantee the faithful performance by the supplier/distributor/manufacturer/contractor/consultant of its obligations under the Contract, I/we shall submit a Performance Securing Declaration within a maximum period of ten (10) calendar days from the receipt of the Notice of Award prior to the signing of the Contract.
2. I/We accept that: I/we will be automatically disqualified from bidding for any procurement contract with any procuring entity for a period of one (1) year for the first offense, or two (2) years **for the second offense**, upon receipt of your Blacklisting Order if I/We have violated my/our obligations under the Contract;
3. I/We understand that this Performance Securing Declaration shall cease to be valid upon:
 - a. issuance by the Procuring Entity of the Certificate of Final Acceptance, subject to the following conditions:
 - i. Procuring Entity has no claims filed against the contract awardee;
 - ii. It has no claims for labor and materials filed against the contractor; and
 - iii. Other terms of the contract; or
 - b. replacement by the winning bidder of the submitted PSD with a performance security in any of the prescribed forms under Section 39.2 of the 2016 revised IRR of RA No. 9184 as required by the end-user.

IN WITNESS WHEREOF, I/We have hereunto set my/our hand/s this ___ day of [month] [year] at [place of execution].

*[Insert NAME OF BIDDER OR ITS
AUTHORIZED REPRESENTATIVE]
[Insert signatory's legal capacity]
Affiant*

[Jurat]

[Format shall be based on the latest Rules on Notarial Practice]

*** Mandatory Provision pursuant to Government Procurement Policy Board (GPPB) Resolution No. 16-2020 dated September 16, 2020**

Statement of Single Largest Completed Contract (SLCC)

Business Name : _____

Business Address : _____

Name of Contract	Date of Contract	Contract Duration	Owner's Name and Address	Description/Kinds of Goods	Amount of Completed Contract/s*	Date of Delivery

Note: This Statement shall be supported with end user's acceptance or official receipt(s) or sales invoice issued for the contract.

*Adjusted by the bidder to current prices using PSA's consumer price index, when necessary, for the purpose of meeting the SLCC requirement.

Submitted by: _____
(Printed Name & Signature)

Designation : _____

Date : _____

** Mandatory Provision pursuant to Government Procurement Policy Board (GPPB) Resolution No. 16-2020 dated September 16, 2020.*

List of All Ongoing Government and Private Contracts including contracts awarded but not yet started

Business Name : _____
Business Address : _____

Name of the Contract	Date of the Contract	Contract Duration	Owner's Name and Address	Description/Kinds of Goods	Amount of Contract/s	Date of Delivery

Submitted by: _____
(Printed Name & Signature)
Designation : _____
Date : _____

** Mandatory Provision pursuant to Government Procurement Policy Board (GPPB) Resolution No. 16-2020 dated September 16, 2020.*

NFCC COMPUTATION FOR ELIGIBILITY CHECK (REVISED)

A.

		Year 20____
1.	Approved Budget for the Contract (ABC)	
2.	Current Assets	
3.	Current Liabilities	
4.	Net Working Capital	
5.	Total value of all outstanding or ongoing Projects	

B. The Net Financial Contracting Capacity (NFCC) based on the above data is computed as follows:

NFCC= [(Current assets minus current liabilities) (15)] minus the value of all outstanding or uncompleted portions of the projects under ongoing contracts, including awarded contracts yet to be started, coinciding with the contract to be bid.

NFCC=P_____

Submitted by:

Name of Supplier/Distributor/Manufacturer

Signature of Authorized Representative

Date:_____

** Mandatory Provision pursuant to Government Procurement Policy Board (GPPB) Resolution No. 16-2020 dated September 16, 2020*